

Certificate of Analysis

ICP/MS Internal Standard

Catalog Number: IMS-112

Lot Number: CS-1286

Lot Issue Date: 03/19/2018



Expiration Date: 04/30/2025

This Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system. The analyte concentrations were verified by our ISO 17025 accredited laboratory to be within $\pm 2.5\%$, when compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.

Analyte	True Value	Analytical Method	NIST SRM
indium	100.0 \pm 0.5 $\mu\text{g/mL}$	ICP/ ICP-MS	3124

Matrix: 2% nitric acid in low TOC water (< 50 ppb)

ULTRA uses purified acids, 18 megohm double deionized water, calibrated Class A glassware & meticulously cleaned bottles in the manufacturing of ULTRAggrade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001.


John Russo
President
Monica Bourgeois
Director of QA/RA