

**ICP-MS Calibration Standard #5**  
**Mercury at 10 µg/mL**

**Catalog Number: IMS-121**  
**Lot Number: R01235**  
**Job Number: J00017275**  
**Lot Issue Date: 11/21/2013**  
**Expiration Date: 12/31/2017**

This Certified Reference Material (CRM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system. The analyte concentration(s) were prepared and verified by an ISO Guide 34 / ISO 17025 accredited laboratory, and compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.

Analyte	True Value	Analytical Method	NIST SRM
* mercury	10.00 ± 0.05 µg/mL	ICP / ICP-MS	3133

**Matrix:** 5% nitric acid in low TOC water (< 50 ppb)

\* *light sensitive*

ULTRA uses purified acids, 18 megaohm double deionized water, calibrated Class A glassware, and meticulously cleaned bottles in the manufacturing of ULTRAgrade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001.



ISO 17025:2005  
Accredited  
A2LA  
Cert. No. 0851.01

ISO 9001:2000  
Registered  
TUV USA, Inc.  
Cert. No. 06-1004



William J. Leary  
Quality Assurance Manager