

# Certificate of Analysis

**Mix Anioni (7 analiti)**

**Catalog Number: ITA-20**

**Lot Number: CP-5700**

**Lot Issue Date: 11/16/2016**


**Expiration Date: 12/31/2018**

This Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system. The analyte concentration(s) were prepared and verified by an ISO Guide 34 / ISO 17025 accredited laboratory, and compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.


Analyte	True Value				Analytical Method	NIST SRM
fluoride	50.1	±	0.3	µg/mL	IC	second source
chloride	100.1	±	0.5	µg/mL	IC	3182
nitrate (as N)	50.1	±	0.3	µg/mL	IC	3185
phosphate (as P) (inorganic)	50.0	±	0.3	µg/mL	IC	3186
sulfate	100.5	±	0.5	µg/mL	IC	3181

**Matrix:** low TOC water (< 50 ppb)

ULTRA uses purified acids, 18 megaohm double deionized water, calibrated Class A glassware, and meticulously cleaned bottles in the manufacturing of ULTRAGrade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.



John Russo  
President



Monica Bourgeois  
Director of QA/RA