

Pesticides Mixture #2

Product Number NPM-102

Page: 1 of 1

Lot Number: CM-2226

Lot Issue Date: 05-May-2015

Expiration Date: 31-May-2018

This certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
atrazine	001912-24-9	RM03843	1005 ± 5 µg/mL
diphenamid	000957-51-7	RM02895	1000 ± 5 µg/mL
EPTC	000759-94-4	RM02957	1000 ± 5 µg/mL
ethoprop	013194-48-4	RM09752	1005 ± 5 µg/mL
mevinphos	007786-34-7	RM07008	1004 ± 5 µg/mL
prometryn	007287-19-6	NT01791	1001 ± 5 µg/mL
propazine	000139-40-2	RM07383	1003 ± 5 µg/mL
terbutryn(e)	000886-50-0	RM03940	1001 ± 5 µg/mL
triadimefon	043121-43-3	RM03376	1001 ± 5 µg/mL

Matrix: methyl tert-butyl ether (MTBE)**Storage:** Store at Room Temperature (15-30°C).

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.