

Certificate of Analysis

Pesticides Mixture #3

Product Number: NPM-103

Page: 1 of 1

Lot Number: CF-4636

Lot Issue Date: Nov-2009

Expiration Date: Dec-2012

This Certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
butachlor	023184-66-9	261-132B	1005 ± 5 µg/mL
carboxin	005234-68-4	183-76A	1004 ± 5 µg/mL
diazinon	000333-41-5	261-18A	1004 ± 5 µg/mL
metolachlor	051218-45-2	422-111B	1004 ± 5 µg/mL
metribuzin	021087-64-9	238-33A	1005 ± 5 µg/mL
MGK-264 (mixed, total)	000113-48-4	249-46B	1005 ± 5 µg/mL
norflurazon	027314-13-2	121-72A	1004 ± 5 µg/mL
terbufos	013071-79-9	241-93A	1004 ± 5 µg/mL
vernolate	001929-77-7	259-81C	1004 ± 5 µg/mL

Matrix: methyl tert-butyl ether (MTBE)


ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025
Accredited
A2LA
Cert. No. 0851-01

ISO 9001:2000
Registered
TUV USA, Inc.
Cert. No. 06-1004

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 295-2330
www.ultrasci.com



William J. Leary
Quality Assurance Manager