



Certificate of Analysis

Pesticides Mixture #4

Product Number: NPM-104A

Page: 1 of 1

Lot Number: CK-0141A

Lot Issue Date: 27-Jan-2014

Expiration Date: 29-Feb-2016

This certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
alachlor	015972-60-8	RM05776	1002 ± 5 µg/mL
atraton	001610-17-9	RM00679	1001 ± 5 µg/mL
bromacil	000314-40-9	RM03909	999.9 ± 5.0 µg/mL
butylate	002008-41-5	RM02684	1004 ± 5 µg/mL
chlorpropham	000101-21-3	RM02736	999.9 ± 5.0 µg/mL
hexazinone	051235-04-2	RM03124	1003 ± 5 µg/mL
molinate	002212-67-1	RM05642	1002 ± 5 µg/mL
pronamide	023950-58-5	RM03337	1001 ± 5 µg/mL
tetrachlorvinphos	022248-79-9	RM04839	1002 ± 5 µg/mL
tricyclazole	041814-78-2	RM03380	1001 ± 5 µg/mL

Matrix: acetone

Storage: Store at Room Temperature (15-30°C).

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025:2005
Accredited
A2LA
Cert. No. 0851-01

ISO 9001:2008
Registered
TUV USA, Inc.
Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 295-2330
www.ultrasci.com


William J. Leary
Quality Assurance Manager