



Certificate of Analysis

Pesticides Mixture #4

Product Number: NPM-104A

Page: 1 of 1

Lot Number: CS-0238

Lot Issue Date: 19-Jan-2018

Expiration Date: 28-Feb-2020

This ISO Guide 34 Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
alachlor	015972-60-8	RM10831	1003 ± 5 µg/mL
atraton	001610-17-9	RM00679	1004 ± 5 µg/mL
bromacil	000314-40-9	RM11790	1000 ± 5 µg/mL
butylate	002008-41-5	RM04558	997.7 ± 5.0 µg/mL
chlorpropham	000101-21-3	RM03671	1003 ± 5 µg/mL
hexazinone	051235-04-2	RM13246	1002 ± 5 µg/mL
molinate	002212-67-1	RM07614	1004 ± 5 µg/mL
pronamide	023950-58-5	RM11321	1002 ± 5 µg/mL
tetrachlorvinphos	022248-79-9	RM13198	1003 ± 5 µg/mL
tricyclazole	041814-78-2	RM03379	1004 ± 5 µg/mL

Matrix: acetone


Storage: Store at Room Temperature (15° to 30°C).

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 9001
Registered
TUV USA, Inc.


John Russo
President


Monica Bourgeois
Director of QA/RA