

Pesticides Mixture #5

Product Number: NPM-105

Page: 1 of 1

Lot Number: CG-4394

Lot Issue Date: 22-Dec-2010

Expiration Date: 31-Jan-2012

This Certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
dichlorvos	000062-73-7	SZE9140X	1000 $\pm$ 5 μ g/mL
fenarimol	060168-88-9	SZE8042X	1004 $\pm$ 5 μ g/mL
fluridone	059756-60-4	174-125B	1000 $\pm$ 5 μ g/mL
napropamide	015299-99-7	190-54A	1002 $\pm$ 5 μ g/mL
pebulate	001114-71-2	320-103A	1001 $\pm$ 5 μ g/mL
simetryn	001014-70-6	60127	1002 $\pm$ 5 μ g/mL
tebuthiuron	034014-18-1	179-119A	1000 $\pm$ 5 μ g/mL
terbacil	005902-51-2	281-78A	1002 $\pm$ 5 μ g/mL

Matrix: methyl tert-butyl ether (MTBE)**Storage:** Store at Room Temperature (18-25° C)


ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025:2005
Accredited
A2LA
Cert. No. 0851-01

ISO 9001:2008
Registered
TUV USA, Inc.
Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 295-2330
www.ultrasci.com


William J. Leary
Quality Assurance Manager