



Certificate of Analysis

Pesticides Mixture #5

Product Number: NPM-105

Page: 1 of 1

Lot Number: CJ-4182

Lot Issue Date: 04-Dec-2012

Expiration Date: 31-Dec-2013

This certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
dichlorvos	000062-73-7	RM06268	1004 ± 5 µg/mL
fenarimol	060168-88-9	RM02969	1001 ± 5 µg/mL
fluridone	059756-60-4	RM04564	999.9 ± 5.0 µg/mL
napropamide	015299-99-7	RM03187	999.9 ± 5.0 µg/mL
pebulate	001114-71-2	RM03264	1003 ± 5 µg/mL
simetryn	001014-70-6	RM05399	1004 ± 5 µg/mL
tebuthiuron	034014-18-1	RM00247	999.6 ± 5.0 µg/mL
terbacil	005902-51-2	RM03374	1002 ± 5 µg/mL

Matrix: methyl tert-butyl ether (MTBE)

Storage: Store at Room Temperature (18-25° C)

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025:2005
Accredited
A2LA
Cert. No. 0851-01

ISO 9001:2008
Registered
TUV USA, Inc.
Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 295-2330
www.ultrasci.com

William J. Leary
Quality Assurance Manager