

Certificate of Analysis

Pesticides Mixture #5

Product Number NPM-105

Page: 1 of 1

Lot Number: CM-2748

Lot Issue Date: 19-May-2015

Expiration Date: 30-Jun-2016

This certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
dichlorvos	000062-73-7	RM10132	1005 ± 5 µg/mL
fenarimol	060168-88-9	RM07337	1003 ± 5 µg/mL
fluridone	059756-60-4	RM07476	1001 ± 5 µg/mL
napropamide	015299-99-7	RM03746	1003 ± 5 µg/mL
pebulate	001114-71-2	RM03264	1005 ± 5 µg/mL
simetryn	001014-70-6	RM05399	1004 ± 5 µg/mL
tebuthiuron	034014-18-1	RM00247	1002 ± 5 µg/mL
terbacil	005902-51-2	RM03373	1002 ± 5 µg/mL

Matrix: methyl tert-butyl ether (MTBE)

Storage: Store at Room Temperature (15-30°C).

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.