



Certificate of Analysis

Pesticides Mixture #6

Product Number: NPM-106

Page: 1 of 1

Lot Number: CH-3268A

Lot Issue Date: 09-Oct-2014

Expiration Date: 30-Nov-2017

This certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
propachlor	001918-16-7	380-26B	1002 ± 5 µg/mL
trifluralin	001582-09-8	229-132B	1002 ± 5 µg/mL
benefin	001861-40-1	RM00683	1002 ± 5 µg/mL
profluralin	026399-36-0	RM03256	1002 ± 5 µg/mL
isopropalin	033820-53-0	RM03100	1002 ± 5 µg/mL
pendimethalin	040487-42-1	244-30A	1002 ± 5 µg/mL
oxadiazon	019666-30-9	RM03171	1002 ± 5 µg/mL
oxyfluorfen	042874-03-3	RM03174	1002 ± 5 µg/mL

Matrix: methyl tert-butyl ether (MTBE)

Storage: Store at Room Temperature (15-30°C).

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025:2005
Accredited
A2LA
Cert. No. 0851-01

ISO 9001:2008
Registered
TUV USA, Inc.
Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 295-2330
www.ultrasci.com


William J. Leary
Quality Assurance Manager