

Triazine Pesticides Mixture

Product Number: NPM-619**Page:** 1 of 1**Lot Number:** CG-3245**Lot Issue Date:** Sep-2010**Expiration Date:** Oct-2013

This Certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
ametryn	000834-12-8	137-137B	100.0 ± 0.5 µg/mL
atraton	001610-17-9	290-30B	100.5 ± 0.5 µg/mL
atrazine	001912-24-9	29-114A	100.0 ± 0.5 µg/mL
prometon	001610-18-0	338-120B	100.5 ± 0.5 µg/mL
prometryn	007287-19-6	225-45A	100.0 ± 0.5 µg/mL
propazine	000139-40-2	356-40A	100.0 ± 0.5 µg/mL
sebumeton	026259-45-0	13330	100.0 ± 0.5 µg/mL
simetryn	001014-70-6	60127	100.5 ± 0.5 µg/mL
simazine	000122-34-9	395-19B	100.5 ± 0.5 µg/mL
terbuthylazine	005915-41-3	NO4W	100.5 ± 0.5 µg/mL
terbutryn(e)	000886-50-0	356-62B	100.5 ± 0.5 µg/mL

Matrix: acetone**Storage:** Store at Room Temperature (18-25° C)


ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025:2005
Accredited
A2LA
Cert. No. 0851-01

ISO 9001:2008
Registered
TUV USA, Inc.
Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 295-2330
www.ultrasci.com


William J. Leary
Quality Assurance Manager