

Pesticides Mixture

Product Number: PPM-017**Page:** 1 of 1**Lot Number:** CF-1317**Lot Issue Date:** Mar-2009**Expiration Date:** Apr-2012

This Certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
atrazine	001912-24-9	401-7A	50.0 ± 0.3 µg/mL
chlorotoluron	015545-48-9	322-42B	50.0 ± 0.3 µg/mL
cyanazine	021725-46-2	310-107B	50.0 ± 0.3 µg/mL
atrazine-desethyl	006190-65-4	60117	50.0 ± 0.3 µg/mL
diuron	000330-54-1	185-37A	50.0 ± 0.3 µg/mL
hexazinone	051235-04-2	146-17B	50.0 ± 0.3 µg/mL
isoproturon	034123-59-6	364-85B	50.0 ± 0.3 µg/mL
linuron	000330-55-2	190-92B	50.0 ± 0.3 µg/mL
metazachlor	067129-08-2	23350	50.0 ± 0.3 µg/mL
methabenzthiazuron	018691-97-9	90216	50.0 ± 0.3 µg/mL
metobromuron	003060-89-7	129-8A	50.0 ± 0.3 µg/mL
metolachlor	051218-45-2	410-81B	50.0 ± 0.3 µg/mL
metoxuron	019937-59-8	4110X	50.0 ± 0.3 µg/mL
monolinuron	001746-81-2	93540	50.0 ± 0.3 µg/mL
sebuthylazine	007286-69-3	21460	50.0 ± 0.3 µg/mL
simazine	000122-34-9	395-19B	50.0 ± 0.3 µg/mL
terbuthylazine	005915-41-3	N04W	50.0 ± 0.3 µg/mL

Matrix: acetonitrile**Storage:** Store at Room Temperature (18-25° C)


ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025:2005
Accredited
A2LA
Cert. No. 0851-01

ISO 9001:2008
Registered
TUV USA, Inc.
Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 295-2330
www.ultrasci.com


William J. Leary
Quality Assurance Manager