



Certificate of Analysis

Organochlorine Pesticides Mixture

Product Number: PPM-505D

Page: 1 of 1

Lot Number: CJ-1105

Lot Issue Date: 04-Apr-2012

Expiration Date: 30-Apr-2013

This certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
alachlor	015972-60-8	364-34A	50.1 ± 0.3 µg/mL
aldrin	000309-00-2	ER012606-03	20.0 ± 0.1 µg/mL
atrazine	001912-24-9	RM03843-01	500.0 ± 2.5 µg/mL
gamma-BHC	000058-89-9	NT01899	20.0 ± 0.1 µg/mL
dieldrin	000060-57-1	ER030105-03	20.1 ± 0.1 µg/mL
endrin	000072-20-8	35091-18	20.1 ± 0.1 µg/mL
heptachlor	000076-44-8	34750-08	20.1 ± 0.1 µg/mL
heptachlor epoxide - isomer B	001024-57-3	ER090105-05	20.1 ± 0.1 µg/mL
hexachlorobenzene	000118-74-1	DG-04416KY	10.0 ± 0.1 µg/mL
hexachlorocyclopentadiene	000077-47-4	03907AU	20.1 ± 0.1 µg/mL
methoxychlor	000072-43-5	RM01712-08	200.0 ± 1.0 µg/mL
simazine	000122-34-9	RM00622-01	100.0 ± 0.5 µg/mL

Matrix: acetone

Storage: Store at Room Temperature (18-25° C)


ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025:2005
Accredited
A2LA
Cert. No. 0851-01

ISO 9001:2008
Registered
TUV USA, Inc.
Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 295-2330
www.ultrasci.com


William J. Leary
Quality Assurance Manager