

Organochlorine Pesticides Mixture

Product Number: PPM-505E

Page: 1 of 1

Lot Number: CE-3453

Lot Issue Date: Oct-2008

Expiration Date: Oct-2011

This Certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
alachlor	015972-60-8	371-64A	10.0 ± 0.1 µg/mL
atrazine	001912-24-9	401-7A	250.0 ± 1.3 µg/mL
simazine	000122-34-9	395-19B	251.0 ± 1.3 µg/mL
methoxychlor	000072-43-5	357-25C	5.01 ± 0.03 µg/mL
aldrin	000309-00-2	ER012606-03	1.00 ± 0.01 µg/mL
gamma-BHC	000058-89-9	NT01899	1.00 ± 0.01 µg/mL
alpha-chlordane	005103-71-9	27455-76	1.00 ± 0.01 µg/mL
gamma-chlordane	005103-74-2	ER061906-04	1.00 ± 0.01 µg/mL
dieldrin	000060-57-1	ER030105-03	1.00 ± 0.01 µg/mL
endrin	000072-20-8	256-27A	1.00 ± 0.01 µg/mL
heptachlor	000076-44-8	33379-48	1.00 ± 0.01 µg/mL
heptachlor epoxide - isomer B	001024-57-3	ER090105-05	1.00 ± 0.01 µg/mL
hexachlorobenzene	000118-74-1	DG-04416KY	1.00 ± 0.01 µg/mL
hexachlorocyclopentadiene	000077-47-4	03907AU	1.00 ± 0.01 µg/mL
cis-nonachlor	005103-73-1	MAR-21985-83	1.00 ± 0.01 µg/mL
trans-nonachlor	039765-80-5	31532-58	1.00 ± 0.01 µg/mL

Matrix: acetone

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025
Accredited
A2LA
Cert. No. 0851-01

ISO 9001:2000
Registered
TUV USA, Inc.
Cert. No. 06-1004

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 295-2330
www.ultrasci.com



William J. Leary
Quality Assurance Manager