

IRSA 5090 Pesticide Mixture

Product Number: PPM-5090

Page: 1 of 2

Lot Number: CH-2281A

Lot Issue Date: 01-Mar-2012

Expiration Date: 30-Apr-2013

This certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
aldrin	000309-00-2	431-88A	100.0 ± 0.5 µg/mL
alpha-BHC	000319-84-6	27275-15	100.0 ± 0.5 µg/mL
beta-BHC	000319-85-7	ER052506-01	100.0 ± 0.5 µg/mL
gamma-BHC	000058-89-9	NT01899	100.0 ± 0.5 µg/mL
delta-BHC	000319-86-8	ER051805-01	100.0 ± 0.5 µg/mL
alpha-chlordane	005103-71-9	27455-76	100.0 ± 0.5 µg/mL
gamma-chlordane	005103-74-2	ER061906-04	100.0 ± 0.5 µg/mL
4,4'-DDD	000072-54-8	W800	100.0 ± 0.5 µg/mL
4,4'-DDE	000072-55-9	X102	100.0 ± 0.5 µg/mL
4,4'-DDT	000050-29-3	02230KY	100.0 ± 0.5 µg/mL
dieldrin	000060-57-1	ER030105-03	100.0 ± 0.5 µg/mL
endrin	000072-20-8	256-27A	100.0 ± 0.5 µg/mL
alachlor	015972-60-8	RM01478	100.0 ± 0.5 µg/mL
atrazine	001912-24-9	RM01560	100.0 ± 0.5 µg/mL
2,4'-DDD	000053-19-0	18307EA	100.0 ± 0.5 µg/mL
2,4'-DDE	003424-82-6	ER122203-02	100.0 ± 0.5 µg/mL
2,4'-DDT	000789-02-6	AJO-22599-59	100.0 ± 0.5 µg/mL
isodrin	000465-73-6	EX-02001LK	100.0 ± 0.5 µg/mL
hexachlorobenzene	000118-74-1	DG04416KY	100.0 ± 0.5 µg/mL
heptachlor	000076-44-8	E01H	100.0 ± 0.5 µg/mL


ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025
Accredited
A2LA
Cert. No. 0851-01

ISO 9001
Registered
TUV USA, Inc.
Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 401-295-2330
www.ultrasci.com


 William J. Leary
 Quality Assurance Manager

IRSA 5090 Pesticide Mixture

Product Number: PPM-5090

Page: 2 of 2

Lot Number: CH-2281A

Lot Issue Date: 01-Mar-2012

Expiration Date: 30-Apr-2013

Analyte	CAS#	Analyte Lot	True Value
heptachlor epoxide - isomer B	001024-57-3	ER090105-05	100.0 ± 0.5 µg/mL
endosulfan I	000959-98-8	ER012105-02	100.0 ± 0.5 µg/mL
endosulfan II	033213-65-9	ER080805-01	100.0 ± 0.5 µg/mL
methoxychlor	000072-43-5	410-71B	100.0 ± 0.5 µg/mL
pentachlorobenzene	000608-93-5	RM00911	100.0 ± 0.5 µg/mL

Matrix: acetone

Storage: Store at Room Temperature (18-25°C). Do Not Refrigerate.


ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025
Accredited
A2LA
Cert. No. 0851-01

ISO 9001
Registered
TUV USA, Inc.
Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 401-295-2330
www.ultrasci.com


William J. Leary
Quality Assurance Manager