

Organochlorine Pesticides Mixture

Product Number: PPM-525C

Page: 1 of 1

Lot Number: CG-0318

Lot Issue Date: Jan-2010

Expiration Date: Feb-2013

This Certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
alachlor	015972-60-8	4232X	100.0 ± 0.5 µg/mL
aldrin	000309-00-2	ER012606-03	100.0 ± 0.5 µg/mL
atrazine	001912-24-9	29-114A	100.0 ± 0.5 µg/mL
alpha-chlordane	005103-71-9	27455-76	100.0 ± 0.5 µg/mL
gamma-chlordane	005103-74-2	ER061906-04	100.0 ± 0.5 µg/mL
gamma-BHC	000058-89-9	NT01899	100.0 ± 0.5 µg/mL
endrin	000072-20-8	256-27A	100.0 ± 0.5 µg/mL
heptachlor	000076-44-8	33379-48	100.0 ± 0.5 µg/mL
heptachlor epoxide - isomer B	001024-57-3	ER090105-05	100.0 ± 0.5 µg/mL
methoxychlor	000072-43-5	224-108B	100.0 ± 0.5 µg/mL
trans-nonachlor	039765-80-5	31532-58	100.0 ± 0.5 µg/mL
simazine	000122-34-9	395-19B	100.0 ± 0.5 µg/mL

Matrix: acetone**Storage:** Store at Room Temperature (18-25°C). Do Not Refrigerate.


ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025:2005
Accredited
A2LA
Cert. No. 0851-01

ISO 9001:2008
Registered
TUV USA, Inc.
Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 295-2330
www.ultrasci.com


William J. Leary
Quality Assurance Manager