

Organochlorine Pesticide Mixture

Product Number: PPM-525F

Page: 1 of 1

Lot Number: CH-0328

Lot Issue Date: 03-Feb-2011

Expiration Date: 28-Feb-2013

This Certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
alachlor	015972-60-8	SZE9139X	100.0 ± 0.5 µg/mL
aldrin	000309-00-2	431-88A	100.0 ± 0.5 µg/mL
atrazine	001912-24-9	440-88C	100.0 ± 0.5 µg/mL
dieldrin	000060-57-1	ER030105-03	100.0 ± 0.5 µg/mL
endrin	000072-20-8	35091-18	100.0 ± 0.5 µg/mL
heptachlor	000076-44-8	224-124B	100.0 ± 0.5 µg/mL
heptachlor epoxide - isomer B	001024-57-3	ER090105-05	100.0 ± 0.5 µg/mL
hexachlorobenzene	000118-74-1	DG-04416KY	100.0 ± 0.5 µg/mL
hexachlorocyclopentadiene	000077-47-4	4878	100.0 ± 0.5 µg/mL
gamma-BHC	000058-89-9	NT01899	100.0 ± 0.5 µg/mL
methoxychlor	000072-43-5	410-71B	100.0 ± 0.5 µg/mL
propachlor	001918-16-7	235-132A	100.0 ± 0.5 µg/mL
simazine	000122-34-9	440-103B	100.0 ± 0.5 µg/mL
trifluralin	001582-09-8	229-132B	100.0 ± 0.5 µg/mL

Matrix: acetone**Storage:** Store at Room Temperature (18-25° C)


ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025:2005
Accredited
A2LA
Cert. No. 0851-01

ISO 9001:2008
Registered
TUV USA, Inc.
Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 295-2330
www.ultrasci.com


William J. Leary
Quality Assurance Manager