

Organochlorine Pesticide Mixture

Product Number PPM-525F

Page: 1 of 1

Lot Number: CM-1886

Lot Issue Date: 20-Apr-2015

Expiration Date: 31-May-2017

This certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
alachlor	015972-60-8	RM09791	100.3 ± 0.5 µg/mL
aldrin	000309-00-2	RM00635	100.3 ± 0.5 µg/mL
atrazine	001912-24-9	RM03843	100.2 ± 0.5 µg/mL
dieldrin	000060-57-1	RM03957	100.4 ± 0.5 µg/mL
endrin	000072-20-8	NT01979	100.1 ± 0.5 µg/mL
heptachlor	000076-44-8	RM03111	100.1 ± 0.5 µg/mL
heptachlor epoxide - isomer B	001024-57-3	RM05373	100.1 ± 0.5 µg/mL
hexachlorobenzene	000118-74-1	RM03125	100.3 ± 0.5 µg/mL
hexachlorocyclopentadiene	000077-47-4	RM07487	100.4 ± 0.5 µg/mL
gamma-BHC	000058-89-9	RM02694	100.3 ± 0.5 µg/mL
methoxychlor	000072-43-5	RM09243	100.3 ± 0.5 µg/mL
propachlor	001918-16-7	RM06880	100.2 ± 0.5 µg/mL
simazine	000122-34-9	RM00622	100.3 ± 0.5 µg/mL
trifluralin	001582-09-8	RM07745	100.3 ± 0.5 µg/mL

Matrix: acetone**Storage:** Store at Room Temperature (15-30°C).

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.