

Certificate of Analysis

ISO Guide 34

Organochlorine Pesticides Standard

Product Number: PPM-608B

Page: 1 of 2

Lot Number: CS-4945

Lot Issue Date: 24-Oct-2018

Expiration Date: 31-May-2019

This ISO Guide 34 Reference Material (RM) was manufactured and verified in accordance with Agilent's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
endosulfan II	033213-65-9	RM07039	20.1 ± 0.1 µg/mL
endosulfan sulfate	001031-07-8	RM12593	20.1 ± 0.1 µg/mL
endrin	000072-20-8	RM00557	20.0 ± 0.1 µg/mL
endrin aldehyde	007421-93-4	RM12540	20.0 ± 0.1 µg/mL
heptachlor	000076-44-8	RM03111	20.0 ± 0.1 µg/mL
aldrin	000309-00-2	RM00635	20.0 ± 0.1 µg/mL
alpha-BHC	000319-84-6	RM13413	20.1 ± 0.1 µg/mL
beta-BHC	000319-85-7	RM13194	20.0 ± 0.1 µg/mL
delta-BHC	000319-86-8	RM01531	20.0 ± 0.1 µg/mL
gamma-BHC	000058-89-9	RM02694	20.0 ± 0.1 µg/mL
4,4'-DDD	000072-54-8	RM00615	20.0 ± 0.1 µg/mL
4,4'-DDE	000072-55-9	RM02892	20.1 ± 0.1 µg/mL
4,4'-DDT	000050-29-3	RM06888	20.1 ± 0.1 µg/mL
dieldrin	000060-57-1	RM13088	20.0 ± 0.1 µg/mL
endosulfan I	000959-98-8	RM00834	20.1 ± 0.1 µg/mL
heptachlor epoxide - isomer B	001024-57-3	RM00308	20.1 ± 0.1 µg/mL

Certificate of Analysis ISO Guide 34

Organochlorine Pesticides Standard

Product Number: PPM-608B

Page: 2 of 2

Lot Number: CS-4945

Lot Issue Date: 24-Oct-2018

Expiration Date: 31-May-2019

This ISO Guide 34 Reference Material (RM) was manufactured and verified in accordance with Agilent's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Matrix: methanol (methyl alcohol)

Storage: Store Refrigerated (2° to 8°C), Do Not Freeze.

Agilent uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.


Monica Bourgeois
QMS Representative