

Organochlorine Pesticides Mixture

Product Number: PPM-608C

Page: 1 of 1

Lot Number: CK-0316

Lot Issue Date: 07-Feb-2013

Expiration Date: 28-Feb-2014

This certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
aldrin	000309-00-2	RM00635	20.1 ± 0.1 µg/mL
alpha-BHC	000319-84-6	RM02725	20.0 ± 0.1 µg/mL
beta-BHC	000319-85-7	RM02724	20.1 ± 0.1 µg/mL
delta-BHC	000319-86-8	RM05696	20.0 ± 0.1 µg/mL
gamma-BHC	000058-89-9	RM02694	20.0 ± 0.1 µg/mL
4,4'-DDD	000072-54-8	RM00615	100.4 ± 0.5 µg/mL
4,4'-DDE	000072-55-9	RM02891	20.1 ± 0.1 µg/mL
4,4'-DDT	000050-29-3	RM02881	100.2 ± 0.5 µg/mL
dieldrin	000060-57-1	RM03957	20.1 ± 0.1 µg/mL
endosulfan I	000959-98-8	RM05767	20.0 ± 0.1 µg/mL
endosulfan II	033213-65-9	RM02905	100.4 ± 0.5 µg/mL
endosulfan sulfate	001031-07-8	NT00714	100.5 ± 0.5 µg/mL
endrin	000072-20-8	RM00557	100.3 ± 0.5 µg/mL
endrin aldehyde	007421-93-4	RM04131	20.1 ± 0.1 µg/mL
heptachlor	000076-44-8	RM03111	20.1 ± 0.1 µg/mL
heptachlor epoxide - isomer B	001024-57-3	RM05373	20.1 ± 0.1 µg/mL
methoxychlor	000072-43-5	NT01497	20.1 ± 0.1 µg/mL

Matrix: methanol (methyl alcohol)**Storage:** Store at Room Temperature (15-30°C). Do Not Refrigerate.


ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025
Accredited
A2LA
Cert. No. 0851-01

ISO 9001
Registered
TUV USA, Inc.
Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 401-295-2330
www.ultrasci.com


William J. Leary
Quality Assurance Manager