

Certificate of Analysis

ISO Guide 34

Organochlorine Pesticides Standard

Product Number: PPM-608C

Page: 1 of 2

Lot Number: CS-4004

Lot Issue Date: 03-Aug-2018

Expiration Date: 30-Sep-2019

This ISO Guide 34 Reference Material (RM) was manufactured and verified in accordance with Agilent's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
aldrin	000309-00-2	RM00635	20.1 ± 0.1 µg/mL
alpha-BHC	000319-84-6	RM14073	20.1 ± 0.1 µg/mL
beta-BHC	000319-85-7	RM13194	20.1 ± 0.1 µg/mL
delta-BHC	000319-86-8	RM09156	20.0 ± 0.1 µg/mL
gamma-BHC	000058-89-9	RM02694	20.1 ± 0.1 µg/mL
4,4'-DDD	000072-54-8	RM00615	100.1 ± 0.5 µg/mL
4,4'-DDE	000072-55-9	RM02891	20.0 ± 0.1 µg/mL
4,4'-DDT	000050-29-3	RM06888	100.4 ± 0.5 µg/mL
dieldrin	000060-57-1	RM03957	20.1 ± 0.1 µg/mL
endosulfan I	000959-98-8	RM00834	20.0 ± 0.1 µg/mL
endosulfan II	033213-65-9	RM07039	99.9 ± 0.5 µg/mL
endosulfan sulfate	001031-07-8	RM06741	100.0 ± 0.5 µg/mL
endrin	000072-20-8	RM00557	100.4 ± 0.5 µg/mL
endrin aldehyde	007421-93-4	RM04131	20.0 ± 0.1 µg/mL
heptachlor	000076-44-8	RM03111	20.1 ± 0.1 µg/mL
heptachlor epoxide - isomer B	001024-57-3	RM00308	20.0 ± 0.1 µg/mL
methoxychlor	000072-43-5	RM12528	20.0 ± 0.1 µg/mL

Certificate of Analysis ISO Guide 34

Organochlorine Pesticides Standard

Product Number: PPM-608C

Page: 2 of 2

Lot Number: CS-4004

Lot Issue Date: 03-Aug-2018


Expiration Date: 30-Sep-2019

This ISO Guide 34 Reference Material (RM) was manufactured and verified in accordance with Agilent's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Matrix: methanol (methyl alcohol)

Storage: Store at Room Temperature (15° to 30°C), Do Not Refrigerate.

Agilent uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.


Monica Bourgeois
QMS Representative