



Certificate of Analysis

Pesticides Extractables Mixture

Product Number: PPM-625B

Page: 1 of 1

Lot Number: CM-3400

Lot Issue Date: 25-Jun-2015

Expiration Date: 31-Jul-2017

This certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
aldrin	000309-00-2	RM00635	20.1 ± 0.1 µg/mL
beta-BHC	000319-85-7	RM09584	20.1 ± 0.1 µg/mL
delta-BHC	000319-86-8	RM09156	20.0 ± 0.1 µg/mL
4,4'-DDD	000072-54-8	RM00615	20.1 ± 0.1 µg/mL
4,4'-DDE	000072-55-9	RM02891	20.0 ± 0.1 µg/mL
4,4'-DDT	000050-29-3	RM06888	20.0 ± 0.1 µg/mL
dieldrin	000060-57-1	RM03957	20.1 ± 0.1 µg/mL
endosulfan sulfate	001031-07-8	RM06741	20.1 ± 0.1 µg/mL
endrin aldehyde	007421-93-4	RM04131	20.0 ± 0.1 µg/mL
heptachlor	000076-44-8	RM03111	20.1 ± 0.1 µg/mL
heptachlor epoxide - isomer B	001024-57-3	RM05373	20.1 ± 0.1 µg/mL

Matrix: methanol (methyl alcohol)

Storage: Store at Room Temperature (15-30°C). Do Not Refrigerate.

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025:2005
Accredited
A2LA
Cert. No. 0851-01

ISO 9001:2008
Registered
TUV USA, Inc.
Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 295-2330
www.ultrasci.com

William J. Leary
Quality Assurance Manager