



Certificate of Analysis

Organochlorine Pesticides Mixture

Product Number: PPM-808B

Page: 1 of 1

Lot Number: CL-0129

Lot Issue Date: 24-Jan-2014

Expiration Date: 29-Feb-2016

This certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
aldrin	000309-00-2	RM00635	251.0 ± 1.3 µg/mL
alpha-BHC	000319-84-6	RM02725	250.0 ± 1.3 µg/mL
beta-BHC	000319-85-7	NT01501	250.1 ± 1.3 µg/mL
delta-BHC	000319-86-8	RM05696	250.1 ± 1.3 µg/mL
gamma-BHC	000058-89-9	RM02694	250.0 ± 1.3 µg/mL
4,4'-DDD	000072-54-8	RM00615	250.7 ± 1.3 µg/mL
4,4'-DDE	000072-55-9	RM02891	250.5 ± 1.3 µg/mL
4,4'-DDT	000050-29-3	RM02881	249.8 ± 1.2 µg/mL
dieldrin	000060-57-1	RM03957	249.8 ± 1.2 µg/mL
endosulfan I	000959-98-8	RM00834	250.7 ± 1.3 µg/mL
endosulfan II	033213-65-9	RM07039	250.0 ± 1.3 µg/mL
endosulfan sulfate	001031-07-8	RM06741	250.4 ± 1.3 µg/mL
endrin	000072-20-8	RM04132	250.2 ± 1.3 µg/mL
endrin aldehyde	007421-93-4	RM04131	249.9 ± 1.2 µg/mL
heptachlor	000076-44-8	RM03111	250.0 ± 1.3 µg/mL
heptachlor epoxide - isomer B	001024-57-3	RM05373	250.9 ± 1.3 µg/mL
methoxychlor	000072-43-5	RM06490	1002 ± 5 µg/mL

Matrix: hexane/toluene (1:1)

Storage: Store at Room Temperature (15-30°C). Do Not Refrigerate.

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025:2005
Accredited
A2LA
Cert. No. 0851-01

ISO 9001:2008
Registered
TUV USA, Inc.
Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 295-2330
www.ultrasci.com

William J. Leary
Quality Assurance Manager