

Certificate of Analysis

Organochlorine Pesticides Mixture

Product Number: PPM-808C**Page:** 1 of 2**Lot Number:** CJ-3745**Lot Issue Date:** 18-Oct-2012**Expiration Date:** 30-Nov-2014

This certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
aldrin	000309-00-2	RM00635	1005 ± 5 µg/mL
alpha-BHC	000319-84-6	RM02725	1005 ± 5 µg/mL
beta-BHC	000319-85-7	RM02724	1004 ± 5 µg/mL
delta-BHC	000319-86-8	RM02723	1002 ± 5 µg/mL
gamma-BHC	000058-89-9	RM02694	1003 ± 5 µg/mL
alpha-chlordane	005103-71-9	RM02742	1002 ± 5 µg/mL
gamma-chlordane	005103-74-2	RM02726	1005 ± 5 µg/mL
4,4'-DDD	000072-54-8	RM00615	1003 ± 5 µg/mL
4,4'-DDE	000072-55-9	RM02892	1004 ± 5 µg/mL
4,4'-DDT	000050-29-3	RM02881	1004 ± 5 µg/mL
dieldrin	000060-57-1	RM03957	1004 ± 5 µg/mL
endosulfan I	000959-98-8	RM00834	1004 ± 5 µg/mL
endosulfan II	033213-65-9	RM02905	1004 ± 5 µg/mL
endosulfan sulfate	001031-07-8	NT00714	1004 ± 5 µg/mL
endrin	000072-20-8	RM00557	1003 ± 5 µg/mL
endrin aldehyde	007421-93-4	RM04131	1004 ± 5 µg/mL
endrin ketone	053494-70-5	RM00212	1005 ± 5 µg/mL
heptachlor	000076-44-8	RM03111	1004 ± 5 µg/mL
heptachlor epoxide - isomer B	001024-57-3	RM05373	1002 ± 5 µg/mL
methoxychlor	000072-43-5	RM04555	1005 ± 5 µg/mL

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025
Accredited
A2LA
Cert. No. 0851-01

ISO 9001
Registered
TUV USA, Inc.
Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 401-295-2330
www.ultrasci.com


William J. Leary
Quality Assurance Manager

Certificate of Analysis

Organochlorine Pesticides Mixture

Product Number: PPM-808C

Page: 2 of 2

Lot Number: CJ-3745

Lot Issue Date: 18-Oct-2012

Expiration Date: 30-Nov-2014

Analyte

CAS#

Analyte Lot

True Value

Matrix: hexane/toluene (1:1)

Storage: Store at Room Temperature (15-30°C). Do Not Refrigerate.

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025
Accredited
A2LA
Cert. No. 0851-01

ISO 9001
Registered
TUV USA, Inc.
Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 401-295-2330
www.ultrasci.com


William J. Leary
Quality Assurance Manager