



Certificate of Analysis

Organochlorine Pesticides Mixture

Product Number: PPM-808F

Page: 1 of 1

Lot Number: CK-1260A

Lot Issue Date: 14-May-2014

Expiration Date: 30-Jun-2016

This certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
chlorobenzilate	000510-15-6	NT00526	1003 ± 5 µg/mL
diallate (total)	002303-16-4	RM06219	1004 ± 5 µg/mL
1,2-dibromo-3-chloropropane	000096-12-8	RM03703	1003 ± 5 µg/mL
hexachlorobenzene	000118-74-1	RM03125	1004 ± 5 µg/mL
hexachlorocyclopentadiene	000077-47-4	RM02368	1004 ± 5 µg/mL
isodrin	000465-73-6	RM03076	1003 ± 5 µg/mL

Matrix: hexane/toluene (1:1)

Storage: Store at Room Temperature (15-30°C).

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025:2005
Accredited
A2LA
Cert. No. 0851-01

ISO 9001:2008
Registered
TUV USA, Inc.
Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 295-2330
www.ultrasci.com

William J. Leary
Quality Assurance Manager