

Organochlorine Pesticides Mixture

Product Number: PPM-808G**Page:** 1 of 1**Lot Number:** CF-2864**Lot Issue Date:** Sep-2009**Expiration Date:** Aug-2012

This Certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
alachlor	015972-60-8	90427	1002 ± 5 µg/mL
captafol	002425-06-1	361-22B	1000 ± 5 µg/mL
chloroneb	002675-77-6	196-39A	1002 ± 5 µg/mL
chloropropylate	005836-10-2	112-131A	1000 ± 5 µg/mL
chlorothalonil	001897-45-6	265-114A	1002 ± 5 µg/mL
DCPA	001861-32-1	229-1B	1004 ± 5 µg/mL
dichlone	000117-80-6	191-26B	1004 ± 5 µg/mL
dicofol	000115-32-2	418-66A	1002 ± 5 µg/mL
etridiazole	002593-15-9	122-66A	1002 ± 5 µg/mL
mirex	002385-85-5	35038-81	1004 ± 5 µg/mL
nitrofen	001836-75-5	341-131A	1004 ± 5 µg/mL
trans-nonachlor	039765-80-5	31532-58	1002 ± 5 µg/mL
pentachloronitrobenzene	000082-68-8	MTC87615	1004 ± 5 µg/mL
permethrins (mixed isomers,total)	052645-53-1	225-100B	1002 ± 5 µg/mL
perthane	000072-56-0	302-112A	1004 ± 5 µg/mL
propachlor	001918-16-7	380-26B	1004 ± 5 µg/mL
trifluralin	001582-09-8	249-94A	1000 ± 5 µg/mL

Matrix: hexane/toluene (1:1)**Storage:** Store at Room Temperature (18-25° C)


ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025:2005
Accredited
A2LA
Cert. No. 0851-01

ISO 9001:2008
Registered
TUV USA, Inc.
Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 295-2330
www.ultrasci.com


William J. Leary
Quality Assurance Manager