

Reference Material Certificate
Product Information Sheet

Product Name: GC/LC Pesticide Standard #3

Lot Number: 0006743861

Product Number: PSM-100-AC

Lot Issue Date: 16-May-2023

Storage Conditions: Store Refrigerated (2° - 8°C).

Expiration Date: 30-Jun-2025

Component Name	Concentration		Uncertainty	CAS#	Analyte Lot
benzoylprop ethyl	100.2	±	0.5 µg/mL	022212-55-1	RM02677
bromoxynil	100.1	±	0.5 µg/mL	001689-84-5	RM19138
crimidine	100.2	±	0.5 µg/mL	000535-89-7	RM13161
cycloate	100.4	±	0.5 µg/mL	001134-23-2	RM16258
cycloxydim	100.4	±	0.5 µg/mL	101205-02-1	RM17003
desmedipham	100.1	±	0.5 µg/mL	013684-56-5	RM18519
desmetryne	100.4	±	0.5 µg/mL	001014-69-3	RM18027
diclofop methyl	100.4	±	0.5 µg/mL	051338-27-3	RM21199
dimoxystrobin	100.3	±	0.5 µg/mL	149961-52-4	RM13935
fenpropathrin	100.1	±	0.5 µg/mL	039515-41-8	RM18367
flufenacet	100.4	±	0.5 µg/mL	142459-58-3	RM13363
flutriafol	100.3	±	0.5 µg/mL	076674-21-0	RM19865
folpet	100.3	±	0.5 µg/mL	000133-07-3	RM09175
fuberidazole	100.2	±	0.5 µg/mL	003878-19-1	RM18116
isopropalin	100.5	±	0.5 µg/mL	033820-53-0	RM03705
mefenpyr-diethyl	100.4	±	0.5 µg/mL	135590-91-9	RM14855
metolachlor	100.2	±	0.5 µg/mL	051218-45-2	RM20722
myclobutanil	100.1	±	0.5 µg/mL	088671-89-0	RM20197
picoxystrobin	100.2	±	0.5 µg/mL	117428-22-5	RM16652
propargite	100.2	±	0.5 µg/mL	002312-35-8	RM19066
pyrimethanil	100.1	±	0.5 µg/mL	053112-28-0	RM18632
tralkoxydim	100.2	±	0.5 µg/mL	087820-88-0	RM16622

Matrix: acetonitrile

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:

This analytical reference standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:

The certification of this analytical reference standard is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:



Monica Bourgeois
QMS Representative