

Reference Material Certificate
Product Information Sheet

Product Name: FDA Pesticide Standard GC no. 2
Product Number: PSM-101-B
Storage Conditions: Store at Room Temperature (15° to 30°C).

Lot Number: 0006726285
Lot Issue Date: 01-Feb-2023
Expiration Date: 31-Mar-2024

Component Name	Concentration	Uncertainty	CAS#	Analyte Lot
2-phenylphenol	100.4 ±	0.5 µg/mL	000090-43-7	RM09146
bifenthrin	100.2 ±	0.5 µg/mL	082657-04-3	RM19152
DCPA	100.3 ±	0.5 µg/mL	001861-32-1	RM12906
diphenylamine	100.4 ±	0.5 µg/mL	000122-39-4	RM12958
endosulfan I	100.2 ±	0.5 µg/mL	000959-98-8	RM20883
endosulfan II	100.4 ±	0.5 µg/mL	033213-65-9	RM18940
etofenprox	100.3 ±	0.5 µg/mL	080844-07-1	RM18181
fenpropathrin	100.3 ±	0.5 µg/mL	039515-41-8	RM16597
fenvalerate	100.1 ±	0.5 µg/mL	051630-58-1	RM10549
hexachlorobenzene	100.3 ±	0.5 µg/mL	000118-74-1	RM03125
iprodione	100.2 ±	0.5 µg/mL	036734-19-7	RM21016
lambda-cyhalothrin	100.4 ±	0.5 µg/mL	091465-08-6	RM18268
tetradifon	100.2 ±	0.5 µg/mL	000116-29-0	RM16748
tolclofos-methyl	100.2 ±	0.5 µg/mL	057018-04-9	RM15666
tolyfluanid	100.2 ±	0.5 µg/mL	000731-27-1	RM16911
trans-chlordane	100.3 ±	0.5 µg/mL	005103-74-2	RM02726
trans-permethrin	100.3 ±	0.5 µg/mL	061949-77-7	RM16650

Matrix: acetone

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material (RM) standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above. Purity values are taken from approved vendor raw material certificates.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference (RM) standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:

This analytical reference (RM) standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:

The certification of this analytical reference standard (RM) is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:



Monica Bourgeois
QMS Representative