

Reference Material Certificate
Product Information Sheet

Product Name: FDA Pesticide Standard GC no. 5

Lot Number: 0006743279

Product Number: PSM-101-E

Lot Issue Date: 16-May-2023

Storage Conditions: Store at Room Temperature (15° to 30°C).

Expiration Date: 30-Jun-2024

Component Name	Concentration	Uncertainty	CAS#	Analyte Lot
2,4'-DDE	100.1 ±	0.5 µg/mL	003424-82-6	RM20201
2,6-diisopropyl-naphthalene	100.0 ±	0.5 µg/mL	024157-81-1	RM09901
benfluralin	100.3 ±	0.5 µg/mL	001861-40-1	RM00683
chlorothalonil	100.5 ±	0.5 µg/mL	001897-45-6	RM18302
cycloate	100.5 ±	0.5 µg/mL	001134-23-2	RM16258
cyfluthrin	100.4 ±	0.5 µg/mL	068359-37-5	RM18189
dicofol	100.1 ±	0.5 µg/mL	000115-32-2	RM14743
endosulfan sulfate	100.1 ±	0.5 µg/mL	001031-07-8	RM20849
endrin	100.1 ±	0.5 µg/mL	000072-20-8	RM14459
fenamiphos	100.1 ±	0.5 µg/mL	022224-92-6	RM03736
fipronil	100.0 ±	0.5 µg/mL	120068-37-3	RM20910
fluopicolide	100.0 ±	0.5 µg/mL	239110-15-7	RM16763
heptachlor epoxide - isomer B	100.3 ±	0.5 µg/mL	001024-57-3	RM21250
methoprene	100.0 ±	0.5 µg/mL	040596-69-8	RM19844
prallethrin	100.0 ±	0.5 µg/mL	023031-36-9	RM18276
propanil	100.0 ±	0.5 µg/mL	000709-98-8	RM03261
resmethrin	100.4 ±	0.5 µg/mL	010453-86-8	RM12003
simazine	100.5 ±	0.5 µg/mL	000122-34-9	RM14327
tetramethrin	100.0 ±	0.5 µg/mL	007696-12-0	RM16228

Matrix: acetone

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:

This analytical reference standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.


Expiration of Certification:

The certification of this analytical reference standard is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:



Monica Bourgeois
QMS Representative