

Reference Material Certificate
Product Information Sheet

Product Name: FDA Pesticide Standard LC no. 3

Lot Number: 0006739670

Product Number: PSM-101-H

Lot Issue Date: 18-Apr-2023

Storage Conditions: Store Frozen (-25° to -10°C).

Expiration Date: 31-Aug-2023

Component Name	Concentration	Uncertainty	CAS#	Analyte Lot
aldicarb	100.1 ±	0.5 µg/mL	000116-06-3	RM16389
aldicarb sulfone	100.2 ±	0.5 µg/mL	001646-88-4	RM15172-01
carbetamide	100.4 ±	0.5 µg/mL	016118-49-3	RM17114
clethodim	100.0 ±	0.5 µg/mL	099129-21-2	RM21059
clofentezine	100.2 ±	0.5 µg/mL	074115-24-5	RM17374
desmedipham	100.0 ±	0.5 µg/mL	013684-56-5	RM18519
fenobucarb	100.1 ±	0.5 µg/mL	003766-81-2	RM17478
fenoxycarb	100.4 ±	0.5 µg/mL	072490-01-8	RM17719
fenuron	100.2 ±	0.5 µg/mL	000101-42-8	RM02972
forchlorfenuron	100.2 ±	0.5 µg/mL	068157-60-8	RM16839
metconazole	100.2 ±	0.5 µg/mL	125116-23-6	RM17856
pencycuron	100.3 ±	0.5 µg/mL	066063-05-6	RM18340
phenmedipham	100.4 ±	0.5 µg/mL	013684-63-4	RM17436
propoxur	100.1 ±	0.5 µg/mL	000114-26-1	RM16246
pymetrozine	100.1 ±	0.5 µg/mL	123312-89-0	RM13625
siduron (mix of isomers)	100.0 ±	0.5 µg/mL	001982-49-6	RM03313
tebuthiuron	100.2 ±	0.5 µg/mL	034014-18-1	RM17522
temephos	100.2 ±	0.5 µg/mL	003383-96-8	RM18486

Matrix: acetonitrile

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:

This analytical reference standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:

The certification of this analytical reference standard is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:



Monica Bourgeois
QMS Representative