



Certificate of Analysis

LC Pesticide Solution #7

Product Number: PSM-105-I

Page: 1 of 1

Lot Number: CR-4617

Lot Issue Date: 10-Oct-2017

Expiration Date: 30-Nov-2019

This ISO Guide 34 Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
amisulbrom	348635-87-0	RM11962	50.0 ± 0.3 µg/mL
benzobicyclon	156963-66-5	RM13078	50.0 ± 0.3 µg/mL
dinotefuran	165252-70-0	RM10018	49.9 ± 0.2 µg/mL
ethaboxam	162650-77-3	RM10950	50.1 ± 0.3 µg/mL
etoxazol	153233-91-1	RM12737	49.9 ± 0.2 µg/mL
flonicamid	158062-67-0	RM12972	50.1 ± 0.3 µg/mL
imicyafos	140163-89-9	RM10044	50.1 ± 0.3 µg/mL
pyrazolynate	058011-68-0	RM13076	50.1 ± 0.3 µg/mL
pyrifluquinazon	337458-27-2	RM10405	50.1 ± 0.3 µg/mL
sethoxydim	074051-80-2	RM12264	49.9 ± 0.2 µg/mL
sulfoxaflor	946578-00-3	RM11908	49.8 ± 0.2 µg/mL
uniconazole	083657-22-1	RM11922	50.2 ± 0.3 µg/mL

Matrix: acetonitrile


Storage: Store Refrigerated (2° - 8°C).

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 9001
Registered
TUV USA, Inc.


John Russo
President


Monica Bourgeois
Director of QA/RA