



Certificate of Analysis

LC Pesticide Solution #8

Product Number: PSM-105-J

Page: 1 of 1

Lot Number: CR-4511

Lot Issue Date: 04-Oct-2017

Expiration Date: 30-Nov-2019

This ISO Guide 34 Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
ametoctradin	865318-97-4	RM13003	50.1 ± 0.3 µg/mL
demeton-s-methyl	000919-86-8	RM10246	50.0 ± 0.3 µg/mL
dimethylvinphos	002274-67-1	RM10099	50.1 ± 0.3 µg/mL
fluacrypyrim	229977-93-9	RM09855	50.0 ± 0.3 µg/mL
fosthiazate	098886-44-3	RM11835	50.1 ± 0.3 µg/mL
imibenconazole	086598-92-7	RM12424	50.1 ± 0.3 µg/mL
inabenfide	082211-24-3	RM12038	50.0 ± 0.3 µg/mL
oxaziclomefone	153197-14-9	RM11911	50.0 ± 0.3 µg/mL
penoxsulam	219714-96-2	RM09021	50.1 ± 0.3 µg/mL
saflufenacil	372137-35-4	RM12288	50.1 ± 0.3 µg/mL
spinetoram	187166-40-1	RM10615	50.1 ± 0.3 µg/mL
tiadinil	223580-51-6	RM08163	50.2 ± 0.3 µg/mL

Matrix: acetonitrile


Storage: Store Refrigerated (2° - 8°C).

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 9001
Registered
TUV USA, Inc.


John Russo
President


Monica Bourgeois
Director of QA/RA