

Certificate of Analysis

ISO Guide 34 Reference Material

Product Number: PSM-105-L
Lot Number: CM-5634

Lot Issue Date: 29-Oct-2015
Expiration Date: 30-Nov-2017

Product Name: LC/GC Pesticide Solution #1

Description:

This Reference Material (RM) was gravimetrically prepared in accordance with ISO Guide 34 and under ULTRA Scientific's ISO 9001 registered quality system. The neat materials used for this product have been verified by ULTRA's ISO 17025 laboratory and under ULTRA's ISO Guide 34 accreditation. The analyte concentrations were verified by ULTRA's ISO 17025 accredited laboratory. For each analyte, the true value, with its uncertainty value calculated at the 95% confidence level, is reported below.

Analyte	CAS#	Analyte Lot	Calculated Value	True Value	Traceability & Method
cyflufenamid	180409-60-3	RM09513	50.2 µg/mL	49.4 ± 0.2 µg/mL	NW06934;LC/DAD
cyprodinil	121552-61-2	RM03665	50.1 µg/mL	49.6 ± 0.1 µg/mL	NW06934;LC/DAD
fenamidone	161326-34-7	RM09950	50.0 µg/mL	48.5 ± 0.2 µg/mL	NW06934;LC/DAD
fenothiocarb	062850-32-2	RM03725	50.2 µg/mL	50.6 ± 1.0 µg/mL	CM-2814;LC/DAD
ferimzone	089269-64-7	RM09848	50.0 µg/mL	51.3 ± 2.3 µg/mL	NW06934;LC/DAD
fluopyram	658066-35-4	RM10451	50.0 µg/mL	50.2 ± 0.1 µg/mL	NW06934;LC/DAD
methidathion	000950-37-8	RM03160	50.2 µg/mL	51.3 ± 0.3 µg/mL	NW06934;LC/DAD
metolachlor	051218-45-2	RM10830	50.2 µg/mL	51.3 ± 1.6 µg/mL	NW06934;LC/DAD
metribuzin	021087-64-9	RM09798	49.9 µg/mL	50.4 ± 0.1 µg/mL	NW06934;LC/DAD
novaluron	116714-46-6	RM10417	50.1 µg/mL	51.3 ± 0.4 µg/mL	NW06934;LC/DAD
penhiopyrad	183675-82-3	RM09969	50.2 µg/mL	50.7 ± 0.3 µg/mL	NW06934;LC/DAD
picoxystrobin	117428-22-5	RM05573	50.0 µg/mL	51.2 ± 0.2 µg/mL	NW06934;LC/DAD
pyridaphenthion	000119-12-0	RM01715	50.1 µg/mL	51.1 ± 0.1 µg/mL	NW06934;LC/DAD
quinoclamine	002797-51-5	RM03330	50.0 µg/mL	49.9 ± 0.2 µg/mL	NW06934;LC/DAD
simeconazole	149508-90-7	RM09887	49.9 µg/mL	50.2 ± 0.4 µg/mL	NW06934;LC/DAD
spiromesifen	283594-90-1	RM09770	50.1 µg/mL	51.1 ± 0.3 µg/mL	NW06934;LC/DAD
thiazopyr	117718-60-2	RM03435	50.2 µg/mL	50.3 ± 0.3 µg/mL	NW06934;LC/DAD


Solvent: Acetonitrile

Storage: Store Refrigerated (2° - 8°C)

ISO Guide 34 Reference Material

Product Number: PSM-105-L
Lot Number: CM-5634

Lot Issue Date: 29-Oct-2015
Expiration Date: 30-Nov-2017



Traceability:

Traceability has been established through an unbroken chain of comparisons, each having stated uncertainties. Comparisons are based on appropriate physical or chemical measurements, including gravimetric or volumetric dilution, where the mass or volume of a solution before and after dilution is measured. The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1, ISO 9001, ISO 17025, and ISO Guide 34. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 819.

Estimation of Uncertainties:

The true value is reported, with its uncertainty value calculated at the 95% confidence level.

Homogeneity:

This RM was formulated and unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Intended Use:

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods and continuing calibration verification.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening and should be processed without delay for the true value to be valid within the stated uncertainties. Each unit contains slightly more than the stated labeled volume to facilitate transfer of the material for testing.

Should crystallization occur after refrigeration, gentle warming (<40°C) and shaking of the container is usually sufficient to re-dissolve the material. If this is unsuccessful, an ultrasonic bath may be used. Solutions containing volatile components (such as gases) should be chilled prior to opening to minimize headspace problems.

Hazards:

Refer to the Safety Data Sheet for information regarding this RM.

Expiration of Certification:

The certification of this RM is valid, within the measurement uncertainty specified, until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Certificate of Analysis

ISO Guide 34 Reference Material

Product Number: PSM-105-L
Lot Number: CM-5634

Lot Issue Date: 29-Oct-2015
Expiration Date: 30-Nov-2017

Maintenance of Certification:

The real-time, long term stability of the RM may be monitored over the lifetime of the certification. If substantive changes occur that affect the certification before the expiration of this certificate, ULTRA Scientific will notify the purchaser.



William J. Leary
Vice President, QA/RA