

# Certificate of Analysis

## Tuning Solution for ICP-MS 7500cs

**Agilent Part Number: 5185-5959**
**Lot Number: 14-53GSX2**

Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity	Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity
Ce	7440-45-1	1.00 µg/L	1.02 µg/L	3110*	Ce(NO <sub>3</sub> ) <sub>2</sub> · 6H <sub>2</sub> O	99.99+	Mg	7439-95-4	1.00 µg/L	1.00 µg/L	3131a*	Mg	99.99+
Co	7440-48-4	1.00 µg/L	1.00 µg/L	3113*	Co	99.99+	Tl	7440-28-0	1.00 µg/L	1.01 µg/L	3158*	TlNO <sub>3</sub>	99.99+
Li	7439-93-2	1.00 µg/L	0.990 µg/L	3129a*	Li <sub>2</sub> CO <sub>3</sub>	99.99+	Y	7440-65-5	1.00 µg/L	0.998 µg/L	3167a*	Y <sub>2</sub> O <sub>3</sub>	99.99+

\* - indicates NIST SRM

† - indicates CRM (when NIST SRM is not available)

**Purity grades:**

Starting Materials: Shown above

Matrix:

 2wt% HNO<sub>3</sub>: HNO<sub>3</sub> (CAS No. 7697-37-2) high purity grade

**Traceability:**

This standard has been produced gravimetrically and volumetrically using ISO 9001 quality procedures. ICP / ICP-MS Spectrometer was used to determine the concentration of the main elements via NIST SRMs shown above, as well as the impurities. Other reference standards used: all 8, 14-28GS.

**Trace Metallic Impurities in the Actual Solution, in µg/L, via ICP-MS Analysis, results are accurate to ±10%:**


Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.
Ag	<0.01	Cs	<0.01	Ho	<0.01	Nd	<0.01	Ru	<0.01	Te	<0.01
Al	<0.03	Cu	<0.01	In	<0.01	Ni	<0.01	Sb	<0.01	Th	<0.01
As	<0.01	Dy	<0.01	Ir	<0.01	P	<10	Sc	<0.01	Tl	<0.02
Au	<0.01	Er	<0.01	K	<0.1	Pb	<0.01	Se	<0.01	Tm	<0.01
B	<0.1	Eu	<0.01	La	<0.01	Pd	<0.01	Si	<5	U	<0.01
Ba	<0.01	Fe	<0.1	Lu	<0.01	Pr	<0.01	Sm	<0.01	V	<0.01
Be	<0.01	Ga	<0.01	Mn	<0.01	Pt	<0.01	Sn	<0.01	W	<0.01
Bi	<0.01	Gd	<0.01	Mo	<0.01	Rb	<0.01	Sr	<0.01	Yb	<0.01
Ca	<0.1	Ge	<0.01	Na	<0.03	Re	<0.01	Ta	<0.01	Zn	<0.03
Cd	<0.01	Hf	<0.01	Nb	<0.01	Rh	<0.01	Tb	<0.01	Zr	<0.01
Cr	<0.01	Hg	<0.01								

Balances are calibrated regularly with weight sets traceable to NIST.

Agilent reference standards are guaranteed stable and accurate to ±5% of measured analyte concentration. For these solutions we use the highest purity acids applicable, 18 megohm double deionized water and acid-leached, triple rinsed bottles. All glassware used is class A.

Date of release: September 15, 2012

Date of expiration: March 31, 2013


  
 QC Coordinator  
 CertiPrep, Inc.