

# Certificate of Analysis

## ICP-MS Stock Tuning Solution (100 mL)

**Agilent Part Number: 5188-6564**
**Lot Number: 1-137YJY2**

Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity	Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity
Ce	7440-45-1	10.0 µg/mL	10.0 µg/mL	3110*	Ce <sub>x</sub> (CO <sub>3</sub> ) <sub>1-x</sub> (H <sub>2</sub> O)	99.99+	Tl	7440-28-0	10.0 µg/mL	10.0 µg/mL	3158*	TlNO <sub>3</sub>	99.99+
Co	7440-48-4	10.0 µg/mL	10.1 µg/mL	3113*	Co	99.99+	Y	7440-65-6	10.0 µg/mL	10.1 µg/mL	3167a*	Y <sub>2</sub> O <sub>3</sub>	99.99+
Ll	7439-93-2	10.0 µg/mL	10.0 µg/mL	3129a*	Li <sub>2</sub> CO <sub>3</sub>	99.99+							

\* - indicates NIST SRM

† - Indicates CRM (when NIST SRM is not available)

**Purity grades:**

Starting Materials: Shown above

Matrix:

 2% HNO<sub>3</sub>: HNO<sub>3</sub> (CAS No. 7697-37-2) high purity grade

**Traceability:**

This standard has been produced gravimetrically and volumetrically using ISO 9001 quality procedures. Agilent ICP / ICP-MS Spectrometer was used to determine the concentration of the main elements via NIST SRMs shown above, as well as the impurities. Other reference standards used: 3-166MKB, 4-124MKB, 4-58MKB.

Trace Metallic Impurities in the Actual Solution, in µg/L, via Agilent ICP-MS Analysis, results are accurate to ±10%:


Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.
Ag	0.4	Ce	<0.08	Ho	<0.01	Nd	<0.01	Ru	<0.07	Te	<0.1
Al	0.9	Cu	<0.2	In	<0.03	Ni	<0.01	Sb	<0.03	Th	<0.01
As	<0.3	Dy	<0.02	Ir	<0.07	P	<100	Sc	<0.3	Tl	<2
Au	<0.04	Er	<0.7	K	<300	Pb	<0.07	Se	<1	Tm	<0.01
B	<0.8	Eu	<0.01	La	<0.01	Pd	0.04	Si	<100	U	<0.01
Ba	<0.1	Fe	0.8	Lu	<0.01	Pf	0.2	Sm	<0.03	V	<0.1
Be	<0.5	Ga	0.4	Mg	0.3	Pt	<0.05	Sn	<0.03	W	<0.06
Bi	<0.05	Gd	0.1	Mn	<0.5	Rb	<0.06	Sr	<0.04	Yb	<0.02
Ce	<10	Ge	<0.01	Mo	<0.02	Re	<0.01	Ta	<0.01	Zn	0.3
Cd	<0.01	Hf	<0.02	Na	8	Rh	<0.2	Tb	1	Zr	<0.08
Cr	<0.5	Hg	<0.1	Nb	<0.02						

Balances are calibrated regularly with weight sets traceable to NIST.

Agilent reference standards are guaranteed stable and accurate to ±0.5% of measured analyte concentration. This uncertainty is at 95% confidence interval, a coverage factor of 2. For these solutions we use the highest purity acids applicable, 18 megohm double deionized water and acid-leached, triple rinsed bottles. All glassware used is class A. This standard was manufactured following the guidelines set forth under ISO 17025 and ISO Guide 34 regulations.

Date of release: May 31, 2018

Date of expiration: November 30, 2019


  
 QC Coordinator