

Certificate of Analysis

ICP-MS Stock Tuning Solution (100 mL)

Agilent Part Number: 5188-6564
Lot Number: 22-85VYY2

Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity	Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity
Ce	7440-45-1	10.0 µg/mL	10.0 µg/mL	3110*	Ce(NO ₃) ₆ ·6H ₂ O	99.99+	Tl	7440-28-0	10.0 µg/mL	9.90 µg/mL	3158*	TlNO ₃	99.99+
Co	7440-48-4	10.0 µg/mL	9.91 µg/mL	3113*	Co	99.99+	Y	7440-85-5	10.0 µg/mL	10.0 µg/mL	3167a*	Y ₂ O ₃	99.99+
Li	7439-93-2	10.0 µg/mL	9.90 µg/mL	3129a*	Li ₂ CO ₃	99.99+							

* - indicates NIST SRM

† - indicates CRM (when NIST SRM is not available)

Purity grades:

Starting Materials: Shown above

Matrix:

 2% HNO₃: HNO₃ (CAS No. 7697-37-2) high purity grade

Traceability:

This standard has been produced gravimetrically and volumetrically using ISO 9001 quality procedures. Agilent ICP / ICP-MS Spectrometer was used to determine the concentration of the main elements via NIST SRMs shown above, as well as the impurities. Other reference standards used: 21-218VY, 21-113VY.

Trace Metallic Impurities in the Actual Solution, in µg/L, via Agilent ICP-MS Analysis, results are accurate to ±10%:


Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.
Ag	<0.1	Cs	<0.04	Ho	0.2	Nd	<0.05	Ru	<0.2	Te	<0.4
Al	0.8	Cu	<0.03	In	<0.02	Ni	<0.3	Sb	<0.08	Th	<0.08
As	1	Dy	<0.01	Ir	<0.09	P	<100	Sc	<0.4	Ti	<0.01
Au	<0.3	Er	<1	K	<100	Pb	<0.08	Se	<0.4	Tm	<0.04
B	<0.5	Eu	<0.02	La	0.1	Pd	<0.2	Si	<100	U	<0.01
Ba	<0.1	Fe	<1	Lu	<0.02	Pr	0.2	Sm	<0.05	V	<0.2
Be	<0.07	Ga	0.3	Mg	<0.08	Pt	<0.1	Sn	<0.05	W	<0.1
Bi	<0.03	Gd	0.05	Mn	<0.2	Rb	<0.03	Sr	<0.04	Yb	0.1
Ca	<10	Ge	<0.01	Mo	<0.4	Re	<0.04	Ta	<0.04	Zn	<0.08
Cd	<0.01	Hf	<0.03	Na	2	Rh	<0.08	Tb	<0.1	Zr	<0.1
Cr	<0.2	Hg	<2	Nb	<0.07						

Balances are calibrated regularly with weight sets traceable to NIST.

Agilent reference standards are guaranteed stable and accurate to ±0.5% of measured analyte concentration. For these solutions we use the highest purity acids applicable, 18 megohm double deionized water and acid-leached, triple rinsed bottles. All glassware used is class A.

Date of release: August 15, 2015

Date of expiration: February 28, 2017



 QC Coordinator