

Certificate of Analysis

ICP-MS Stock Tuning Solution (100 mL)

Agilent Part Number: 5188-6564
Lot Number: 3-68MKBY2

Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity	Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity
Ce	7440-45-1	10.0 µg/mL	10.1 µg/mL	3110*	Ce(CO ₃) ₂ ·x(H ₂ O)	99.99+	Tl	7440-28-0	10.0 µg/mL	10.1 µg/mL	3158*	TlNO ₃	99.99+
Co	7440-48-4	10.0 µg/mL	10.0 µg/mL	3113*	Co	99.99+	Y	7440-85-5	10.0 µg/mL	10.1 µg/mL	3167a*	Y ₂ O ₃	99.99+
Li	7439-93-2	10.0 µg/mL	10.1 µg/mL	3129e*	Li ₂ CO ₃	99.99+							

* - Indicates NIST SRM

† - Indicates CRM (when NIST SRM is not available)

Purity grades:

Starting Materials: Shown above

Matrix:

 2% HNO₃: HNO₃ (CAS No. 7697-37-2) high purity grade

Traceability:

This standard has been produced gravimetrically and volumetrically using ISO 9001 quality procedures. Agilent ICP / ICP-MS Spectrometer was used to determine the concentration of the main elements via NIST SRMs shown above, as well as the impurities. Other reference standards used: 23-166VY, 22-151VY.

Trace Metallic Impurities in the Actual Solution, in µg/L, via Agilent ICP-MS Analysis, results are accurate to ±10%:


Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.
Ag	<0.02	Ce	<0.1	Ho	<0.01	Nd	<0.03	Ru	<0.03	Te	<0.3
Al	<0.2	Cu	<0.3	In	<0.01	Ni	0.2	Sb	<0.02	Th	<0.04
As	<0.4	Dy	<0.02	Ir	<0.02	P	<100	Sc	<0.3	Tl	<0.01
Au	0.1	Er	<0.01	K	5	Pb	0.2	Se	<6	Tm	<0.01
B	<0.3	Eu	<0.02	La	0.02	Pd	<0.4	Si	<100	U	<0.03
Ba	<0.1	Fe	5	Lu	<0.01	Pr	0.1	Sm	<0.03	V	<0.1
Be	<0.07	Ga	<0.4	Mg	0.5	Pt	<0.1	Sn	<0.1	W	<0.07
Bi	<0.01	Gd	<0.2	Mn	0.3	Rb	<0.1	Sr	<0.05	Yb	<0.03
Ca	<10	Ge	<0.01	Mo	0.02	Re	<0.02	Ta	<0.01	Zn	0.5
Cd	<0.01	Hf	<0.01	Na	2	Rh	<0.2	Tb	0.9	Zr	<0.1
Cr	<0.2	Hg	<0.2	Nb	<0.02						

Balances are calibrated regularly with weight sets traceable to NIST.

Agilent reference standards are guaranteed stable and accurate to ±0.5% of measured analyte concentration. This uncertainty is at 95% confidence interval, a coverage factor of 2. For these solutions we use the highest purity acids applicable, 18 megohm double deionized water and acid-leached, triple rinsed bottles. All glassware used is class A. This standard was manufactured following the guidelines set forth under ISO 17025 and ISO Guide 34 regulations.

Date of release: March 15, 2017

Date of expiration: September 30, 2018



 QC Coordinator