



# Certificate of Analysis

## ICP-MS Tuning Solution 10 ppm 100mL

Agilent Part Number: 5190-0465

Lot Number: 21-168VYY2

Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity	Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity
Ce	7440-45-1	10.0 µg/mL	9.93 µg/mL	3110*	Ce(NO <sub>3</sub> ) <sub>6</sub> ·6H <sub>2</sub> O	99.99+	Mg	7439-95-4	10.0 µg/mL	9.97 µg/mL	3131a*	Mg	99.99+
Co	7440-48-4	10.0 µg/mL	9.99 µg/mL	3113*	Co	99.99+	Tl	7440-28-0	10.0 µg/mL	10.0 µg/mL	3158*	TlNO <sub>3</sub>	99.99+
Li	7439-93-2	10.0 µg/mL	9.97 µg/mL	3129a*	Li <sub>2</sub> CO <sub>3</sub>	99.99+	Y	7440-65-5	10.0 µg/mL	9.97 µg/mL	3167a*	Y <sub>2</sub> O <sub>3</sub>	99.99+

\* - indicates NIST SRM

† - indicates CRM (when NIST SRM is not available)

### Purity grades:

Starting Materials: Shown above  
 Matrix:  
 2% HNO<sub>3</sub>: HNO<sub>3</sub> (CAS No. 7697-37-2) high purity grade

### Traceability:

This standard has been produced gravimetrically and volumetrically using ISO 9001 quality procedures. ICP / ICP-MS Spectrometer was used to determine the concentration of the main elements via NIST SRMs shown above, as well as the impurities. Other reference standards used: all 8, 20-11VY, 19-91VY.

### Trace Metallic Impurities in the Actual Solution, in µg/L, via ICP-MS Analysis, results are accurate to ±10%:

Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.
Ag	<0.08	Cs	<0.01	Ho	<0.01	Nd	<0.2	Ru	<1	Te	<0.4
Al	1	Cu	<0.9	In	<0.04	Ni	1	Sb	<0.2	Th	<0.04
As	2	Dy	<0.01	Ir	<0.07	P	<100	Sc	<0.1	Tl	<0.8
Au	<0.06	Er	<0.02	K	5	Pb	<0.2	Se	<5	Tm	<0.01
B	<0.8	Eu	<0.02	La	<0.03	Pd	<0.3	Si	<200	U	<0.01
Ba	<0.2	Fe	5	Lu	<0.01	Pr	0.08	Sm	<0.05	V	<0.07
Be	<0.7	Ga	0.4	Mn	0.7	Pt	<0.03	Sn	<0.3	W	<0.09
Bi	<0.2	Gd	0.06	Mo	<0.5	Rb	<0.1	Sr	<0.05	Yb	<0.01
Ca	20	Ge	<4	Na	6	Re	<0.02	Ta	<0.1	Zn	2
Cd	<0.2	Hf	<0.02	Nb	<0.1	Rh	<0.06	Tb	0.4	Zr	<0.3
Cr	3	Hg	<0.2								

Balances are calibrated regularly with weight sets traceable to NIST.

Agilent reference standards are guaranteed stable and accurate to ±0.5% of measured analyte concentration. For these solutions we use the highest purity acids applicable, 18 megohm double deionized water and acid-leached, triple rinsed bottles. All glassware used is class A.

Date of release: October 31, 2014

Date of expiration: April 30, 2016

*Brijender S. Tank*  
 QC Coordinator