

Certificate of Analysis

ICP-MS Tuning Solution 10 ppm 100mL

Agilent Part Number: 5190-0465
Lot Number: 21-64VY2

Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity	Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity
Ce	7440-45-1	10.0 mg/L	9.93 mg/L	3110*	Ce(NO ₃) ₂ · 6H ₂ O	99.99+	Mg	7439-95-4	10.0 mg/L	10.0 mg/L	3131a*	Mg	99.99+
Co	7440-48-4	10.0 mg/L	9.92 mg/L	3113*	Co	99.99+	Tl	7440-28-0	10.0 mg/L	10.0 mg/L	3158*	TlNO ₃	99.99+
Li	7439-93-2	10.0 mg/L	10.0 mg/L	3129a*	Li ₂ CO ₃	99.99+	Y	7440-65-5	10.0 mg/L	9.97 mg/L	3167a*	Y ₂ O ₃	99.99+

* - indicates NIST SRM

† - indicates CRM (when NIST SRM is not available)

Purity grades:

Starting Materials: Shown above

Matrix:

 2% HNO₃: HNO₃ (CAS No. 7697-37-2) high purity grade

Traceability:

This standard has been produced gravimetrically and volumetrically using ISO 9001 quality procedures. ICP / ICP-MS Spectrometer was used to determine the concentration of the main elements via NIST SRMs shown above, as well as the impurities. Other reference standards used: 13-165VY, 19-91VY, 17-22VY.

Trace Metallic Impurities in the Actual Solution, in µg/L, via ICP-MS Analysis, results are accurate to ±10%:

Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.
Ag	<0.1	Cs	<0.03	Ho	<0.01	Nd	<0.01	Ru	<0.2	Te	<0.2
Al	<1	Cu	<0.3	In	<0.03	Ni	<1	Sb	<0.09	Th	0.06
As	9	Dy	<0.01	Ir	<0.03	P	<100	Sc	<1	Ti	<0.6
Au	<0.06	Er	<0.03	K	<1	Pb	<0.09	Se	<1	Tm	<0.01
B	<0.8	Eu	<0.04	La	0.4	Pd	<0.1	Si	<200	U	<0.02
Ba	<0.1	Fe	2	Lu	<0.01	Pr	0.2	Sm	<0.05	V	<0.2
Be	<0.3	Ga	0.4	Mn	<0.2	Pt	<0.05	Sn	<0.5	W	<0.7
Bi	<0.02	Gd	0.6	Mo	<0.4	Rb	<0.2	Sr	<0.2	Yb	<0.04
Ca	<20	Ge	<0.4	Na	1	Re	<0.06	Ta	<0.05	Zn	<2
Cd	<0.2	Hf	<0.02	Nb	<0.2	Rh	<0.06	Tb	<0.5	Zr	<0.5
Cr	<2	Hg	<0.3								

Balances are calibrated regularly with weight sets traceable to NIST.

Agilent reference standards are guaranteed stable and accurate to ±0.5% of measured analyte concentration. For these solutions we use the highest purity acids applicable, 18 megohm double deionized water and acid-leached, triple rinsed bottles. All glassware used is class A.

Date of release: June 15, 2014

Date of expiration: December 31, 2015

Brijender S Tonk
 QC Coordinator