

Certificate of Analysis

ICP-MS Tuning Solution 10 ppm 100mL

Agilent Part Number: 5190-0465

Lot Number: 3-05MKBY2

Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity	Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity
Ce	7440-45-1	10.0 µg/mL	10.0 µg/mL	3110*	Ce(CO ₃) ₂ ·x(H ₂ O)	99.99+	Mg	7439-95-4	10.0 µg/mL	10.0 µg/mL	3131a*	Mg	99.99+
Co	7440-48-4	10.0 µg/mL	10.1 µg/mL	3113*	Co	99.99+	Tl	7440-28-0	10.0 µg/mL	10.1 µg/mL	3158*	TlNO ₃	99.99+
Li	7439-93-2	10.0 µg/mL	9.95 µg/mL	3129a*	LiCO ₃	99.99+	Y	7440-65-5	10.0 µg/mL	10.1 µg/mL	3167a*	Y ₂ O ₃	99.99+

* - indicates NIST SRM

† - indicates CRM (when NIST SRM is not available)

Purity grades:

Starting Materials: Shown above

Matrix:

2% HNO₃: HNO₃ (CAS No. 7897-37-2) high purity grade

Traceability:

This standard has been produced gravimetrically and volumetrically using ISO 9001 quality procedures. Agilent ICP / ICP-MS Spectrometer was used to determine the concentration of the main elements via NIST SRMs shown above, as well as the impurities. Other reference standards used: 22-121VY, 20-11VY.

Trace Metallic impurities in the Actual Solution, in µg/L, via Agilent ICP-MS Analysis, results are accurate to ±10%:

Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.
Ag	0.01	Ce	<0.01	Ho	<0.01	Nd	<0.01	Ru	<0.01	Te	<0.05
Al	4	Cu	<0.3	In	<0.01	Ni	<0.1	Sb	0.02	Th	<0.01
As	<0.3	Dy	<0.01	Ir	<0.01	P	<100	Sc	<0.02	Tl	<0.5
Au	<0.01	Er	<0.01	K	<3	Pb	0.05	Se	<0.5	Tm	<0.01
B	<0.2	Eu	<0.01	La	0.01	Pd	0.1	Si	<50	U	<0.01
Ba	0.1	Fe	2	Lu	<0.01	Pr	0.2	Sm	<0.01	V	<0.02
Be	<0.03	Ga	0.4	Mn	0.4	Pt	<0.01	Sr	<0.01	W	<0.01
Bi	<0.01	Gd	0.2	Mo	<0.02	Rb	<0.01	Sr	<0.02	Yb	<0.01
Ce	<10	Ge	<0.01	Ne	3	Re	<0.01	Ta	<0.01	Zn	0.4
Cd	<0.01	Hf	<0.01	Nb	<0.01	Rh	0.02	Tb	1	Zr	<0.1
Cr	<0.2	Hg	<0.01								

Balances are calibrated regularly with weight sets traceable to NIST.

Agilent reference standards are guaranteed stable and accurate to ±0.5% of measured analyte concentration. This uncertainty is at 95% confidence interval, a coverage factor of 2. For these solutions we use the highest purity acids applicable, 18 megohm double deionized water and acid-leached, triple rinsed bottles. All glassware used is class A. This standard was manufactured following the guidelines set forth under ISO 17025 and ISO Guide 34 regulations.

Date of release: November 15, 2016

Date of expiration: May 31, 2018

Brijender S. Tork
 QC Coordinator