

Certificate of Analysis

ICP-MS Tuning Solution 10 ppm 100mL

Agilent Part Number: 5190-0465

Lot Number: 3-150MKBY2

Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity	Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity
Ce	7440-45-1	10.0 µg/mL	10.1 µg/mL	3110*	Ce ₂ (CO ₃) ₃ ·x(H ₂ O)	99.99+	Mg	7439-95-4	10.0 µg/mL	9.94 µg/mL	3131a*	Mg	99.99+
Co	7440-48-4	10.0 µg/mL	10.0 µg/mL	3113*	Co	99.99+	Ti	7440-28-0	10.0 µg/mL	10.0 µg/mL	3158*	TiO ₂	99.99+
Li	7439-93-2	10.0 µg/mL	10.0 µg/mL	3129a*	Li ₂ CO ₃	99.99+	Y	7440-65-6	10.0 µg/mL	10.1 µg/mL	3167a*	Y ₂ O ₃	99.99+

* - indicates NIST SRM

† - indicates CRM (when NIST SRM is not available)

Purity grades:

Starting Materials: Shown above

Matrix:

2% HNO₃: HNO₃ (CAS No. 7697-37-2) high purity grade

Traceability:

This standard has been produced gravimetrically and volumetrically using ISO 9001 quality procedures. Agilent ICP / ICP-MS Spectrometer was used to determine the concentration of the main elements via NIST SRMs shown above, as well as the impurities. Other reference standards used: 22-121VY, 23-98VY, 23-115VY, 22-24VY.

Trace Metallic Impurities in the Actual Solution, in µg/L, via Agilent ICP-MS Analysis, results are accurate to ±10%:

Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.
Ag	<2	Cs	<0.04	Ho	<0.01	Nd	<0.02	Ru	<0.2	Te	<0.3
Al	2	Cu	0.3	In	<0.01	Ni	0.4	Sb	<0.05	Th	<0.02
As	0.9	Dy	<0.01	Ir	<0.04	P	<100	Sc	<0.2	Ti	<0.6
Au	<0.1	Er	<0.01	K	20	Pb	<0.2	Se	6	Tm	<0.01
B	<0.8	Eu	<0.01	La	0.04	Pd	<0.3	Si	<100	U	<0.01
Ba	<0.1	Fe	3	Lu	<0.01	Pr	0.09	Sm	<0.03	V	<0.08
Be	<0.3	Ga	0.8	Mn	0.2	Pt	<0.04	Sr	<0.1	W	<0.2
Bi	<0.01	Gd	0.1	Mo	<0.1	Rb	<0.2	Sr	<0.09	Yb	<0.01
Ce	25	Ge	<2	Na	3	Re	<0.02	Ta	<0.1	Zn	1
Cd	<0.06	Hi	<0.04	Nb	<0.1	Rh	<0.1	Tb	0.4	Zr	<0.9
Cr	0.5	Hg	<0.3								

Balances are calibrated regularly with weight sets traceable to NIST.

Agilent reference standards are guaranteed stable and accurate to ±0.5% of measured analyte concentration. This uncertainty is at 95% confidence interval, a coverage factor of 2. For these solutions we use the highest purity acids applicable, 18 megohm double deionized water and acid-leached, triple rinsed bottles. All glassware used is class A. This standard was manufactured following the guidelines set forth under ISO 17025 and ISO Guide 34 regulations.

Date of release: April 15, 2017

Date of expiration: October 31, 2018

Brijender S Tonk
 QC Coordinator

Visit Agilent on the world wide web: www.agilent.com/chem/supplies