

Certificate of Analysis

Multi Element Calibration Standard 2A - HG

Agilent Part Number: 8500-6940-HG

Lot Number: 9-97HGY2

Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity
Hg	7439-97-8	10.0 µg/mL	10.0 µg/mL	3133*	Hg	99.99*

* - Indicates NIST SRM

† - Indicates CRM (when NIST SRM is not available)

Purity grades:

Starting Materials: Shown above

Matrix:

5% HNO₃: HNO₃ (CAS No. 7697-37-2) high purity grade

Traceability:

This standard has been produced gravimetrically and volumetrically using ISO 9001 quality procedures. Agilent ICP / ICP-MS Spectrometer was used to determine the concentration of the main elements via NIST SRMs shown above, as well as the Impurities.

Trace Metallic Impurities in the Actual Solution, in µg/L, via Agilent ICP-MS Analysis, results are accurate to ±10%:

Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.
Ag	<0.06	Cr	<0.05	Ho	<0.01	Nb	<0.01	Ru	<0.01	Th	<0.01
Al	<0.02	Cs	<0.01	In	<0.01	Nd	<0.01	Sb	<0.02	Tl	<0.02
As	<0.1	Cu	<0.01	Ir	<0.01	Ni	<0.03	Se	<0.01	Tl	<0.1
Au	<10	Dy	<0.01	K	<1	P	<20	Se	<0.2	Ym	<0.01
B	<0.4	Er	<0.01	La	<0.01	Pb	<0.01	Si	<30	U	<0.01
Ba	<0.01	Eu	<0.01	Li	<0.2	Pd	<10	Sm	<0.01	V	<0.1
Be	<0.03	Fe	<1	Lu	<0.01	Pr	<0.01	Sn	<0.01	W	<0.07
Bi	<0.01	Ga	<0.01	Mg	<0.02	Pt	<0.02	Sr	<0.01	Y	<0.01
Ce	<1	Gd	<0.01	Mn	<0.01	Rb	<0.01	Ta	<0.01	Yb	<0.01
Cd	<0.02	Ge	<0.06	Mo	<0.03	Re	<0.01	Tb	<0.01	Zn	<0.05
Ce	<0.01	Hf	<0.01	Na	0.3	Rh	<0.01	Te	<0.06	Zr	<0.02
Co	<0.01										


Balances are calibrated regularly with weight sets traceable to NIST.

Density: 1.024 g/mL @ 20°C

Agilent reference standards are guaranteed stable and accurate to ±1% of measured analyte concentration. For these solutions we use the highest purity acids applicable, 18 megohm double deionized water and acid-leached, triple rinsed bottles. All glassware used is class A. This standard was manufactured following the guidelines set forth under ISO 17025 and ISO Guide 34 regulations.

Date of release: January 15, 2017

Date of expiration: July 31, 2018


QC Coordinator

Visit Agilent on the world wide web: www.agilent.com/chem/supplies