

# Certificate of Analysis

## Wash-Nitric Acid Blank

Agilent Part Number: G1820-60258

Lot Number: 23-137GST2

**Purity grades:**

Matrix:

H<sub>2</sub>O: DI Water (CAS No. 7732-18-5) 18Megohm, double deionized (ASTM Type I)

5wt% HNO<sub>3</sub>: HNO<sub>3</sub> (CAS No. 7697-37-2) high purity grade

**Traceability:**

This standard has been produced gravimetrically and volumetrically using ISO 9001 quality procedures.

Trace Metallic Impurities in the Actual Solution, in µg/L, via Agilent ICP-MS Analysis, results are accurate to ±10%:

| Element | Conc. | Element | Conc. | Element | Conc. | Element | Conc. | Element | Conc. | Element | Conc. |
|---------|-------|---------|-------|---------|-------|---------|-------|---------|-------|---------|-------|
| Ag      | <0.01 | Cr      | <0.01 | Ho      | <0.01 | Nb      | <0.01 | Ru      | <0.01 | Th      | <0.01 |
| Al      | <0.03 | Cs      | <0.01 | In      | <0.01 | Nd      | <0.01 | Sb      | <0.01 | Tl      | <0.01 |
| As      | <0.01 | Cu      | <0.01 | Ir      | <0.01 | Ni      | <0.01 | Sc      | <0.01 | Tl      | <0.01 |
| Au      | <0.01 | Dy      | <0.01 | K       | <0.1  | P       | <10   | Se      | <0.01 | Tm      | <0.01 |
| B       | <0.01 | Er      | <0.01 | La      | <0.01 | Pb      | <0.01 | Si      | <5    | U       | <0.01 |
| Ba      | <0.01 | Eu      | <0.01 | Li      | <0.01 | Pd      | <0.01 | Sm      | <0.01 | V       | <0.01 |
| Be      | <0.01 | Fe      | <0.1  | Lu      | <0.01 | Pr      | <0.01 | Sr      | <0.01 | W       | <0.01 |
| Bi      | <0.01 | Ga      | <0.01 | Mg      | <0.01 | Pt      | <0.01 | Ta      | <0.01 | Y       | <0.01 |
| Ca      | <0.1  | Gd      | <0.01 | Mn      | <0.01 | Pb      | <0.01 | Ta      | <0.01 | Yb      | <0.01 |
| Cd      | <0.01 | Ge      | <0.01 | Mo      | <0.01 | Re      | <0.01 | Tb      | <0.01 | Zn      | <0.03 |
| Ce      | <0.01 | Hf      | <0.01 | Na      | <0.03 | Rh      | <0.01 | Te      | <0.01 | Zr      | <0.01 |
| Co      | <0.01 | Hg      | <0.01 |         |       |         |       |         |       |         |       |

Balances are calibrated regularly with weight sets traceable to NIST.

Agilent reference standards are guaranteed stable and accurate to ±10% of listed impurities shown above. For these solutions we use the highest purity acids applicable, 18 megohm double deionized water and acid-leached, triple rinsed bottles. All glassware used is class A. This standard was manufactured following the guidelines set forth under ISO 17025 and ISO Guide 34 regulations.

Date of release: February 15, 2017

Date of expiration: May 31, 2018

*Brijender S. Tonk*  
 QC Coordinator