

Certificate of Analysis

High Sensitivity Tune

Agilent Part Number: ZHP-21-500

Lot Number: 23-49GSX2

Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity	Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity
Ca	7440-45-1	1.00 µg/L	1.01 µg/L	3110*	Ca(NO ₃) ₂ · 6H ₂ O	99.99+	Tl	7440-28-0	1.00 µg/L	1.02 µg/L	3156*	TlNO ₃	99.99+
Li	7439-93-2	1.00 µg/L	1.01 µg/L	3129a*	Li ₂ CO ₃	99.99+	Y	7440-85-5	1.00 µg/L	1.01 µg/L	3167a*	Y ₂ O ₃	99.99+
Mg	7439-95-4	1.00 µg/L	1.01 µg/L	3131a*	Mg(NO ₃) ₂ · 6H ₂ O	99.99+							

* - Indicates NIST SRM

† - Indicates CRM (when NIST SRM is not available)

Purity grades:

Starting Materials: Shown above

Matrix:

2wt% HNO₃: HNO₃ (CAS No. 7697-37-2) high purity grade

Traceability:

This standard has been produced gravimetrically and volumetrically using ISO 9001 quality procedures. Agilent ICP / ICP-MS Spectrometer was used to determine the concentration of the main elements via NIST SRMs shown above, as well as the impurities. Other reference standards used: all 8; 22-102GS.

Trace Metallic Impurities in the Actual Solution, in µg/L, via Agilent ICP-MS Analysis, results are accurate to ±10%:

Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.
Ag	<0.01	Cr	<0.01	Hg	<0.01	Nd	<0.01	Ru	<0.01	Te	<0.01
Al	<0.03	Cs	<0.01	Ho	<0.01	Ni	<0.01	Sb	<0.01	Th	<0.01
As	<0.01	Cu	<0.01	In	<0.01	P	<10	Sc	<0.01	Tl	<0.01
Au	<0.01	Dy	<0.01	Ir	<0.01	Pb	<0.01	Sa	<0.01	Tm	<0.01
B	<0.01	Er	<0.01	K	<0.1	Pd	<0.01	Si	<5	U	<0.01
Ba	<0.01	Eu	<0.01	La	<0.01	Pr	<0.01	Sm	<0.01	V	<0.01
Ba	<0.01	Fe	<0.1	Lu	<0.01	Pt	<0.01	Sn	<0.01	W	<0.01
Bi	<0.01	Ga	<0.01	Mn	<0.01	Rb	<0.01	Sr	<0.01	Yb	<0.01
Ca	<0.1	Gd	<0.01	Mo	<0.01	Re	<0.01	Ta	<0.01	Zn	<0.03
Cd	<0.01	Ge	<0.01	Na	<0.1	Rh	<0.01	Tb	<0.01	Zr	<0.01
Co	<0.01	Hf	<0.01	Nb	<0.01						

Balances are calibrated regularly with weight sets traceable to NIST.

Agilent reference standards are guaranteed stable and accurate to ±5% of measured analyte concentration. This uncertainty is at 95% confidence interval, a coverage factor of 2. For these solutions we use the highest purity acids applicable, 18 megohm double deionized water and acid-leached, triple rinsed bottles. All glassware used is class A. This standard was manufactured following the guidelines set forth under ISO 17025 and ISO Guide 34 regulations.

Date of release: January 31, 2017

Date of expiration: April 30, 2018

Brijender S. Tonk
 QC Coordinator