

Agilent Technologies

Shipped 6141 EASTON ROAD, BLDG 1 PO BOX 310
From: PLUMSTEADVILLE PA 18949-0310
Phone: 800-331-4953 Fax: 215-766-7226

C E R T I F I C A T E O F A N A L Y S I S

AGILENT TECHNOLOGIES, INC. PROJECT #: 01-32447-001
PO#: 0023072
LOGIC CNTR AMERICAS ITEM #: 0104C400264 110
TWIN SPANS BUSINESS PARK CUST ITEM #: 5190-0521EU
NEW CASTLE DE 19720-4578 DATE: 10Sep2010

CYLINDER #: ST0000141051 ANALYTICAL ACCURACY: +/-2%
FILL PRESSURE: 12755 KPAG PRODUCT EXPIRATION: 09Sep2012
SCOTT LOT#: 025301Z


COMPONENT	REQUESTED GAS		ANALYSIS	
	CONC	MOLES	(MOLES)	
CARBON DIOXIDE	600.	PPM	602.0	PPM
METHANE	5.	PPM	4.99	PPM
NITROUS OXIDE	1.	PPM	1.00	PPM
NITROGEN AGILENT		BALANCE		BALANCE


MASTER CYL.# AAL21385
AGILENT P/N 5190-0521EU

MASTER TRACEABLE TO NIST
WEIGHTS

MANUFACTURED DATE: 10Sep2010

SCOTTY SIZE: 110

APPROVED BY: 
JAKE SENKOW



Agilent Technologies

Shipped 6141 EASTON ROAD, BLDG 1 PO BOX 310
 From: PLUMSTEADVILLE PA 18949-0310
 Phone: 800-331-4953 Fax: 215-766-7226
C E R T I F I C A T E O F A N A L Y S I S

AGILENT TECHNOLOGIES, INC. PROJECT #: 01-32447-001
 PO#: 0023072
 LOGIC CNTR AMERICAS ITEM #: 0104C400264 110
 TWIN SPANS BUSINESS PARK CUST ITEM #: 5190-0521EU
 NEW CASTLE DE 19720-4578 DATE: 10Sep2010

CYLINDER #: ST0000141050 ANALYTICAL ACCURACY: +/-2%
 FILL PRESSURE: 12755 KPAG PRODUCT EXPIRATION: 09Sep2012
 SCOTT LOT#: 025301Z

<u>COMPONENT</u>	<u>REQUESTED GAS</u>		<u>ANALYSIS</u>	
	<u>CONC MOLES</u>		<u>(MOLES)</u>	
CARBON DIOXIDE	600.	PPM	602.0	PPM
METHANE	5.	PPM	4.99	PPM
NITROUS OXIDE	1.	PPM	1.00	PPM
NITROGEN AGILENT		BALANCE		BALANCE

MASTER CYL.# AAL21385
 AGILENT P/N 5190-0521EU

MASTER TRACEABLE TO NIST
 WEIGHTS

MANUFACTURED DATE: 10Sep2010

SCOTTY SIZE: 110

APPROVED BY: *Jake Senkow*
 JAKE SENKOW



Agilent Technologies

Shipped 6141 EASTON ROAD, BLDG 1 PO BOX 310
 From: PLUMSTEADVILLE PA 18949-0310
 Phone: 800-331-4953 Fax: 215-766-7226
C E R T I F I C A T E O F A N A L Y S I S

AGILENT TECHNOLOGIES, INC. PROJECT #: 01-32447-001
 PO#: 0023072
 LOGIC CNTR AMERICAS ITEM #: 0104C400264 110
 TWIN SPANS BUSINESS PARK CUST ITEM #: 5190-0521EU
 NEW CASTLE DE 19720-4578 DATE: 10Sep2010

CYLINDER #: ST0000141049 ANALYTICAL ACCURACY: +/-2%
 FILL PRESSURE: 12755 KPAG PRODUCT EXPIRATION: 09Sep2012
 SCOTT LOT#: 025301Z

COMPONENT	REQUESTED GAS		ANALYSIS	
	CONC	MOLES	(MOLES)	
CARBON DIOXIDE	600.	PPM	602.0	PPM
METHANE	5.	PPM	4.99	PPM
NITROUS OXIDE	1.	PPM	1.00	PPM
NITROGEN AGILENT		BALANCE		BALANCE


MASTER CYL.# AAL21385
 AGILENT P/N 5190-0521EU

MASTER TRACEABLE TO NIST
 WEIGHTS

MANUFACTURED DATE: 10Sep2010

SCOTTY SIZE: 110

APPROVED BY: 
 JAKE SENKOW



Agilent Technologies

Shipped 6141 EASTON ROAD, BLDG 1 PO BOX 310
From: PLUMSTEADVILLE PA 18949-0310
Phone: 800-331-4953 Fax: 215-766-7226

C E R T I F I C A T E O F A N A L Y S I S

AGILENT TECHNOLOGIES, INC. PROJECT #: 01-32447-001
PO#: 0023072
LOGIC CNTR AMERICAS ITEM #: 0104C400264 110
TWIN SPANS BUSINESS PARK CUST ITEM #: 5190-0521EU
NEW CASTLE DE 19720-4578 DATE: 10Sep2010

CYLINDER #: ST0000141048 ANALYTICAL ACCURACY: +/-2%
FILL PRESSURE: 12755 KPAG PRODUCT EXPIRATION: 09Sep2012
SCOTT LOT#: 025301Z

COMPONENT	REQUESTED GAS		ANALYSIS	
	CONC	MOLES	(MOLES)	
CARBON DIOXIDE	600.	PPM	602.0	PPM
METHANE	5.	PPM	4.99	PPM
NITROUS OXIDE	1.	PPM	1.00	PPM
NITROGEN AGILENT		BALANCE		BALANCE


MASTER CYL.# AAL21385
AGILENT P/N 5190-0521EU

MASTER TRACEABLE TO NIST
WEIGHTS

MANUFACTURED DATE: 10Sep2010

SCOTTY SIZE: 110

APPROVED BY: Jake Senkow
JAKE SENKOW



Agilent Technologies

Shipped 6141 EASTON ROAD, BLDG 1 PO BOX 310
 From: PLUMSTEADVILLE PA 18949-0310
 Phone: 800-331-4953 Fax: 215-766-7226
C E R T I F I C A T E O F A N A L Y S I S

AGILENT TECHNOLOGIES, INC. PROJECT #: 01-32447-001
 PO#: 0023072
 LOGIC CNTR AMERICAS ITEM #: 0104C400264 110
 TWIN SPANS BUSINESS PARK CUST ITEM #: 5190-0521EU
 NEW CASTLE DE 19720-4578 DATE: 10Sep2010

CYLINDER #: ST0000141047 ANALYTICAL ACCURACY: +/-2%
 FILL PRESSURE: 12755 KPAG PRODUCT EXPIRATION: 09Sep2012
 SCOTT LOT#: 025301Z


COMPONENT	REQUESTED GAS		ANALYSIS	
	CONC	MOLES	(MOLES)	
CARBON DIOXIDE	600.	PPM	602.0	PPM
METHANE	5.	PPM	4.99	PPM
NITROUS OXIDE	1.	PPM	1.00	PPM
NITROGEN AGILENT		BALANCE		BALANCE

MASTER CYL.# AAL21385
 AGILENT P/N 5190-0521EU

MASTER TRACEABLE TO NIST
 WEIGHTS

MANUFACTURED DATE: 10Sep2010

SCOTTY SIZE: 110

APPROVED BY: 
 JAKE SENKOW