



Certificate of Analysis

Organophosphorous Pesticides Mixture

Product Number: SPM-824

Page: 1 of 2

Lot Number: CM-3639

Lot Issue Date: 09-Jul-2015

Expiration Date: 30-Nov-2015

This certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
guthion	000086-50-0	RM06810	200.5 ± 1.0 µg/mL
bolstar	035400-43-2	RM09392	200.0 ± 1.0 µg/mL
chlorpyrifos	002921-88-2	RM09198	199.8 ± 1.0 µg/mL
coumaphos	000056-72-4	RM09648	200.6 ± 1.0 µg/mL
demeton (total, mixed isomers)	008065-48-3	RM09067	201.0 ± 1.0 µg/mL
diazinon	000333-41-5	RM10440	199.4 ± 1.0 µg/mL
dichlorvos	000062-73-7	RM10132	200.8 ± 1.0 µg/mL
disulfoton	000298-04-4	RM05013	200.3 ± 1.0 µg/mL
ethoprop	013194-48-4	RM10504	199.6 ± 1.0 µg/mL
fensulfothion	000115-90-2	RM07770	200.0 ± 1.0 µg/mL
fenthion	000055-38-9	RM10005	199.9 ± 1.0 µg/mL
merphos	000150-50-5	RM10545	200.5 ± 1.0 µg/mL
methyl parathion	000298-00-0	RM07567	200.6 ± 1.0 µg/mL
mevinphos	007786-34-7	RM09377	200.1 ± 1.0 µg/mL
naled	000300-76-5	RM09352	200.9 ± 1.0 µg/mL
phorate	000298-02-2	RM09649	201.0 ± 1.0 µg/mL
fenchlorphos	000299-84-3	RM09170	199.2 ± 1.0 µg/mL
tetrachlorvinphos	022248-79-9	RM04839	200.8 ± 1.0 µg/mL
tokuthion	034643-46-4	RM06645	199.6 ± 1.0 µg/mL
trichloronate	000327-98-0	RM07348	200.7 ± 1.0 µg/mL

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025:2005
Accredited
A2LA
Cert. No. 0851-01

ISO 9001:2008
Registered
TUV USA, Inc.
Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 295-2330
www.ultrasci.com

William J. Leary
Quality Assurance Manager



Certificate of Analysis

Organophosphorous Pesticides Mixture

Product Number: SPM-824

Page: 2 of 2

Lot Number: CM-3639

Lot Issue Date: 09-Jul-2015

Expiration Date: 30-Nov-2015

This certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Matrix: hexane/acetone (1:1)

Storage: Store at Room Temperature (15-30°C).

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025:2005
Accredited
A2LA
Cert. No. 0851-01

ISO 9001:2008
Registered
TUV USA, Inc.
Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 295-2330
www.ultrasci.com

William J. Leary
Quality Assurance Manager