



Certificate of Analysis

TCLP Pesticides Mixture

Product Number: TCLP-530BN

Page: 1 of 1

Lot Number: CK-0224A

Lot Issue Date: 26-Feb-2016

Expiration Date: 31-Mar-2019

This ISO Guide 34 Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
gamma-BHC	000058-89-9	RM02694	99.8 ± 0.5 µg/mL
heptachlor	000076-44-8	RM03111	100.3 ± 0.5 µg/mL
heptachlor epoxide - isomer B	001024-57-3	RM05373	100.0 ± 0.5 µg/mL
endrin	000072-20-8	RM00557	100.0 ± 0.5 µg/mL
methoxychlor	000072-43-5	NT01497	1004 ± 5 µg/mL

Matrix: methanol (methyl alcohol)

Storage: Store at Room Temperature (15° to 30°C).

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 9001
Registered
TUV USA, Inc.


William J. Leary
Quality Assurance Manager