

Phenols Mixture #2

Product Number: US-117N

Page: 1 of 1

Lot Number: CE-3173Z

Lot Issue Date: 06-Jan-2012

Expiration Date: 28-Feb-2014

This Certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
o-cresol	000095-48-7	09416AA	2007 ± 10 µg/mL
m-cresol	000108-39-4	03517AG	2007 ± 10 µg/mL
p-cresol	000106-44-5	09410PI	2003 ± 10 µg/mL
2,6-dichlorophenol	000087-65-0	MS03518LN	2005 ± 10 µg/mL
dinoseb	000088-85-7	BCD-29550-47	2006 ± 10 µg/mL
hexachlorophene	000070-30-4	49C-0296	2008 ± 10 µg/mL
2,3,4,6-tetrachlorophenol	000058-90-2	ER100206-01	2007 ± 10 µg/mL
2,4,5-trichlorophenol	000095-95-4	07521CO	2007 ± 10 µg/mL

Matrix: methylene chloride (dichloromethane)**Storage:** Store at Room Temperature (18-25°C)

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025
Accredited
A2LA
Cert. No. 0851-01

ISO 9001
Registered
TUV USA, Inc.
Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 401-295-2330
www.ultrasci.com



William J. Leary
Quality Assurance Manager