

Pesticides Mixture

Product Number: US-118

Page: 1 of 1

Lot Number: CG-3565

Lot Issue Date: Oct-2010

Expiration Date: May-2011

This Certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
aramite (total)	000140-57-8	B0508	2008 ± 10 µg/mL
chlorobenzilate	000510-15-6	SZE9070X	2002 ± 10 µg/mL
diallate (total)	002303-16-4	444-15A	2002 ± 10 µg/mL
isodrin	000465-73-6	02001LK	2004 ± 10 µg/mL
kepone	000143-50-0	29494-54	2006 ± 10 µg/mL
pronamide	023950-58-5	383-61A	2006 ± 10 µg/mL

Matrix: methylene chloride (dichloromethane)**Storage:** Store at Room Temperature (18-25° C)


ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025:2005
Accredited
A2LA
Cert. No. 0851-01

ISO 9001:2008
Registered
TUV USA, Inc.
Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 295-2330
www.ultrasci.com


William J. Leary
Quality Assurance Manager